


**STRUCTURAL CEILING PLAN**

SCALE: 1/4" = 1'-0"

CEILING PLAN LEGEND		
SUPPLIED/INSTALLED BY SIEMENS		
SYM	DESCRIPTION	DET
D	DCS RAILS ATTACHED TO "P-1001" UNISTRUT	2
E	DCS CARRIAGE MOVES ALONG LONGITUDINAL RAILS	1
F	RAD. SHIELD RAILS ATTACHED TO "P-1001" UNISTRUT	-
G	RADIATION SHIELD SUPPORT CARRIAGE MOVES ALONG RAILS	-
Z	LONGITUDINAL RAIL SUPPORT MOUNTING POINT BOLTED TO UNISTRUT FRAME	2
SUPPLIED/INSTALLED BY CUSTOMER/CONTRACTOR		
SYM	DESCRIPTION	DET
X	"P-1001" UNISTRUT MOUNTED FLUSH WITH FINISHED CEILING. MUST BE LEVEL WITHIN ±1/8"	3

**NOTE:**  
 ALL STRUCTURAL SUPPORT DETAILS SHOWN ARE SAMPLE DETAILS BASED UPON TYPICAL AND STANDARD BUILDING PRACTICES AND ARE NOT INTENDED AS ACTUAL CONSTRUCTION DETAILS. ALL CONSTRUCTION DETAILS AND SUPPORT CALCULATIONS SHALL BE PREPARED BY A PROFESSIONAL STRUCTURAL ENGINEER AT THE CUSTOMER'S EXPENSE. IN THE EVENT AN EXISTING SUPPORT SYSTEM IS TO BE USED, IT WILL BE THE CUSTOMER'S RESPONSIBILITY TO VERIFY THE INTEGRITY OF THAT SYSTEM.



**3 UNISTRUT DETAIL** SCALE: NONE

CEILING HEIGHT RANGE	RECOMMENDED CEILING HEIGHT
9'-0 1/4" - 10'-6"	9'-6 3/8"

PROJECT MANAGER: RICH D'ISIRI TEL: (207) 712-3225 FAX: (207) 929-3776 EMAIL: RICH.D'ISIRI@SIEMENS.COM		<b>SIEMENS</b>	
<b>MAINE MEDICAL CENTER</b> 22 BRAMHALL ST. PORTLAND, ME 04102 PBC SUITE - ARTIS ZEE MULTI-PURPOSE			
PROJECT #: <b>1102468</b>	SHEET: <b>S-102</b>	DATE: 01/04/12	
DRAWN BY: J. PAOLUCCI		CHECKED:	
SCALE: AS NOTED		REF. # 30152323	

**ATTENTION:**

- THIS DRAWING IS DESIGNED TO CONFORM TO FEATURES AND EQUIPMENT REQUIREMENTS PRESENTED AT THE TIME OF THEIR INSTALLATION. SINCE BOTH THESE FACTORS ARE SUBJECT TO DESIGN MODIFICATION, THEY ARE NOT TO BE USED FOR CONSTRUCTION PURPOSES.  
 - THIS SET OF PLANS REPRESENTS A COMPLETE SET OF DETAILS AND SHOULD NOT BE SEPARATED.  
 - IT IS RECOMMENDED THAT THE SIEMENS DRAWINGS BE INCORPORATED WITH THE CONSTRUCTION DOCUMENTS FOR REFERENCE.  
 - ALL DIMENSIONS SHOWN ON THIS DRAWING ARE FROM FINISHED SURFACES.  
 - THIS DRAWING DOES NOT PROVIDE RADIATION SHIELDING REQUIREMENTS FOR X-RAY AND ASSOCIATED EQUIPMENT. THE CUSTOMER IS RESPONSIBLE FOR CONSULTING WITH A REGISTERED RADIATION PHYSICIST TO SPECIFY RADIATION PROTECTION.

NOTE: THIS DRAWING WAS NOT PREPARED BY SMRT. IT IS PROVIDED FOR COORDINATION AND REFERENCE PURPOSES ONLY.

REFERENCE DOCUMENT - NOT FOR CONSTRUCTION

144 Form Sheet 010, Rev. 019  
Portland, Maine 04104  
Tel: (207) 772-3846  
Fax: (207) 772-1070  
www.smrtinc.com

ARCHITECTURE  
ENGINEERING  
PLANNING  
INTERIOR DESIGN  
COMMISSIONING

**SMRT**

PROJECT NORTH

MAINE MEDICAL CENTER  
HOLDING, RADCU & PBC RENOVATION  
PORTLAND, ME

ISSUED FOR CONSTRUCTION  
02.24.12

CURRENT ISSUE STATUS

REV	DESCRIPTION	DATE

GRAPHIC SCALE:  
0' 1"

SCALE:  
PROJECT MANAGER: KD  
K/DRAWN BY:  
A/E OF RECORD: KD  
PROJECT NO.: 11063  
DATE: 2.24.12

SHEET TITLE:  
VENDOR EQUIPMENT  
DRAWINGS SHEET 5

SHEET No. Q-005