



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: CORRIDOR  
ROOM NUMBER: 6004

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	093412 MVD01	4	<b>MONITOR, VIDEO</b> MXU0299: DISPLAY FLAT PANEL NON-TOUCH PIIC SPACE: SPACE	PHILIPS MEDICAL SYSTEMS MXU0299 866390	N	OF/OI
2	093331 WCH01	4	<b>WALL CHANNEL</b> WALL CHANNEL;SEISMIC;19" SPACE: WALL	PHILIPS MEDICAL SYSTEMS MXU175	N	OF/CI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT CARE STATION (QTY 5)  
ROOM NUMBER: 6005

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	087769 MON01	1	<b>MONITOR, CENTRAL STATION</b> PIIC IX SURVEILLANCE CENTRAL MONITORING;19" TOUCH (WxDxH) :48x16x16 SPACE: COUNTER CM:121.9x40.6x40.6 <b>ELECT:</b> CPU: 115 VAC;60 HZ;5.2 AMPS LCD DISPLAY: 115 VAC;60 HZ;48W TYP. UPS: 115 VAC;60 HZ;5.6 AMPS PRINTER: 117 VAC;60 HZ;5.3 AMPS <b>HVAC:</b> MUST MAINTAIN TEMP AT 62-82 DEGREES F AND NON-CONDENSING RELATIVE HUMIDITY AT 30-60% CPU: 215 BTU/HR UPS: 437 BTU/HR (MAX) DATA CONNECTION REQUIRED, SEE TECH SHEET	PHILIPS MEDICAL SYSTEMS INTELLIVUE INFO CENTER IX	N	OF/OI
2	504528 MON17	1	<b>ALLOWANCE, MONITORING, TELEMETRY</b>	ZZZ - MONITORING	N	OF/OI
3	086723 ANA01	3	<b>LAB ANALYZER, GLUCOSE</b> BLOOD GLUCOSE & KETONE MONITORING;BARCODE READER (WxDxH) :3x2x8 LBS: 1 SPACE: SPACE CM:7.6x5.1x20.3 KG: 0.50 <b>ELECT:</b> 2 AA BATTERIES (ALKALINE, LITHIUM, OR NICAD)	ABBOTT DIAGNOSTICS 71411-70 PRECISION PRO	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, EQUIPMENT (QTY 2)  
ROOM NUMBER: 6011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	006264	1	<b>NOTE:</b>	ZZZ - NOTE: (U.S.)	I	OF/OI
			- NEED TO CONFIRM EXISTING			
2	079188 SHV01	2	<b>SHELVING, WIRE</b> 5 SHELVES;STATIONARY;CHROME;SHELF INLAY (WxDxH) :24x24x74 SPACE: FLOOR CM:61.0x61.0x188.0	INTERMETRO INDUSTRIES CORP. A2424NC/74P A2424NC	I	OF/VI
			- NEED TO CONFIRM EXISTING			
3	004335 CRT28	1	<b>CART, UTILITY</b> 2 SHELF;400 LB CAP;MOLDED PLASTIC;BEIGE (WxDxH) :18x39x34 LBS: 31 SPACE: FLOOR CM:45.7x99.1x86.4 KG: 14.10	RUBBERMAID COMM. PRODUCTS 4500-88BEIG 4500-88	I	OF/OI
4	058498	2	<b>SPHYGMOMANOMETER, ANEROID, MOBILE</b> 5 LEG STAND;CERTIFIED ACCURATE TO 3MMHG;LATEX (WxDxH) :18x18x36 LBS: 12 SPACE: FLOOR CM:45.7x45.7x91.4 KG: 5.50	WELCH ALLYN, INC. 7670-03	I	OF/OI
			- NEED TO CONFIRM EXISTING			
5	072678 DOP01	2	<b>DOPPLER, FLOW DETECTOR</b> NON-DIRECTIONAL POCKET DOPPLER W/BUILT-IN (WxDxH) :18x18x36 LBS: 2 SPACE: FLOOR CM:45.7x45.7x91.4 KG: 0.90	HUNTLEIGH HEALTHCARE D900-P-USA/EZ8 D900 DOPPLEX	I	OF/OI
			<b>ELECT:</b> 9V ALKALINE BATTERY - NEED TO CONFIRM EXISTING			



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, EQUIPMENT (QTY 2)  
ROOM NUMBER: 6011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
6	086490	1	<b>OPHTHALMOSCOPE/OTOSCOPE</b> MOBILE STAND; PANOPTIC OPHTHALMOSCOPE;MACROVIEW (WxDxH) :22x22x54 SPACE: FLOOR CM:55.9x55.9x137.2 <b>ELECT:</b> 100-240V;50/60HZ;.18 AMPS MAX - NEED TO CONFIRM EXISTING	WELCH ALLYN, INC. 77710-82M/ACC GS 777 SERIES	I	OF/OI
7	067194	4	<b>I.V. POLE</b> STAINLESS STEEL;5 LEG;3 IN CASTERS;4 HOOK TOP (WxDxH) :23x6x35 LBS: 32 SPACE: FLOOR CM:58.4x15.2x88.9 KG: 14.50 - NEED TO CONFIRM EXISTING	PRYOR PRODUCTS 135	I	OF/OI
8	088643	4	<b>I.V. INFUSION PUMP</b> GENERAL INFUSION PUMP;SINGLE CHANNEL;POLE CLAMP; (WxDxH) :8x8x6 LBS: 10 SPACE: SPACE CM:20.3x20.3x15.2 KG: 4.50 <b>ELECT:</b> 120V;50/60HZ;35W;1A;1 RECHARGEABLE BATTERY LEAD-ACID 6V;POWER CORD:HOSPITAL-GRADE AC CORD 10 FT LONG W/TRANSPARENT PLUG AND RETAINER PLATE - NEED TO CONFIRM EXISTING	HOSPIRA, INC. 30010 PLUM 360/MEDNET	I	OF/OI
9	087247	4	<b>WALKER, FOLDING, ADULT</b> SIDE WALKER;ALUMINUM;USER WT CAP 300LBS;HT ADJUST (WxDxH) :19x16x36 LBS: 4 SPACE: FLOOR CM:48.3x40.6x91.4 KG: 1.60 - NEED TO CONFIRM EXISTING	PERFORMANCE HEALTHCARE 081561752 HEMI DAYS INVACARE	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, EQUIPMENT (QTY 2)  
ROOM NUMBER: 6011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
10	066897	4	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0 - NEED TO CONFIRM EXISTING	MEDLINE INDUSTRIES, INC. MDS89668XW	I	OF/OI
11	096810 BRD01	2	<b>BOARD, PATIENT TRANSFER</b> 529 LB CAPACITY;FOR USE IN RADIOLOGY,SURGERY,ICU (WxDxH) :20x70x2 LBS: 8 SPACE: WALL CM:50.8x177.8x5.1 KG: 3.80 - NEED TO CONFIRM EXISTING	SAMARIT 440.0100 HIGHTEC	I	OF/OI
12	078455 BRD03	2	<b>BOARD, PATIENT TRANSFER</b> SURGIBOARD 47X16";TO AND FROM OR TABLE (WxDxH) :16x47x2 SPACE: WALL CM:40.6x119.4x5.1 - NEED TO CONFIRM EXISTING	MCAULEY MEDICAL INC 440.0400 SAMARIT	I	OF/OI
13	055330	1	<b>LIFTER, PATIENT</b> INFLATABLE;NYLON OXFORD;TESTED UPT 1100LBS (WxDxH) :39x70x30 SPACE: FLOOR CM:99.1x177.8x76.2 <b>ELECT:</b> 120V/60HZ/1100 WATTS;GROUNDED;14' POWER CORD - NEED TO CONFIRM EXISTING	HOVERTECH INTERNATIONAL HJ3902/AIR400G/ACC HOVERJACK	I	OF/OI
14	057580 LFT02	1	<b>CART</b> HOLDS HOVERJACK AIR SUPPLY AND LIFT (WxDxH) :25x18x38 LBS: 26 SPACE: FLOOR CM:63.5x45.7x96.5 KG: 11.80 - NEED TO CONFIRM EXISTING	HOVERTECH INTERNATIONAL HJC-100 HOVER JACK	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, EQUIPMENT (QTY 2)  
ROOM NUMBER: 6011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
15	073144	1	<b>TRANSILLUMINATOR</b> VEIN LOCATOR;INFRARED LIGHT;HAND HELD;BATTERY (WxDxH) :24x24x36 SPACE: FLOOR CM:61.0x61.0x91.4 - NEED TO CONFIRM EXISTING	BODIFLOW, LLC AV300 ACCUVEIN	I	OF/OI
16	094909 CRT07	1	<b>CART, UTILITY</b> MULTIPURPOSE CART;800LB CAPACITY;5"WHEELS (WxDxH) :21x31x42 LBS: 44 SPACE: FLOOR CM:53.3x78.7x106.7 KG: 20.00 - NEED TO CONFIRM EXISTING	VERSACART EZTOTE 390 101-390	I	OF/OI
17	063659	4	<b>TABLE, OVERBED</b> SINGLE TOP;W/VANITY;LOW PROFILE U-SHAPED BASE (WxDxH) :21x32x45 LBS: 53 SPACE: FLOOR CM:53.3x81.3x114.3 KG: 24.10 SPECIFY FINISH SPECIFY FINISH: BASE SPECIFY FINISH: EDGE - NEED TO CONFIRM EXISTING	STRYKER CORP/MEDICAL DIV 3150-000-200 TRU-FIT	I	OF/OI
18	082590	2	<b>WALKER</b> 4-WHEEL WALKER;8" NON-MARRING CASTERS;300LB CAP (WxDxH) :24x26x37 LBS: 20 SPACE: FLOOR CM:61.0x66.0x94.0 KG: 9.10 - NEED TO CONFIRM EXISTING	PERFORMANCE HEALTHCARE 081576974 ROLLATOR WALKER	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, EQUIPMENT (QTY 2)  
ROOM NUMBER: 6011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
19	091438	1	<b>CHAIR, SHOWER</b> 500LB USER WT CAP;14-19" HT ADJ SEAT;W/KNOCK DOWN (WxDxH) :18x20x30 LBS: 8 SPACE: FLOOR CM:45.7x50.8x76.2 KG: 3.40 - NEED TO CONFIRM EXISTING	DRIVE MEDICAL 12021KD-1 DELUXE BARI	I	OF/OI
20	070331 SCL01	2	<b>SCALE, STAND-ON</b> PORTABLE;DIGITAL READOUT;HANDRAIL/HEIGHT GAUGE (WxDxH) :26x26x78 LBS: 62 SPACE: FLOOR CM:66.0x66.0x198.1 KG: 28.10 <b>ELECT:</b> 120V; 6-D BATTERIES - NEED TO CONFIRM EXISTING	SCALE-TRONIX DIV. OF WELCH ALLYN 5002/845010/845233	I	OF/OI
21	046403 IMU01	2	<b>IMAGING, ULTRASOUND SCANNER</b> BLADDER ULTRASOUND SCANNER; W/ROLLING CART (WxDxH) :15x15x30 LBS: 20 SPACE: FLOOR CM:38.1x38.1x76.2 KG: 9.10 <b>ELECT:</b> BATTERY CHARGER:100-120V;60HZ;INPUT CURRENT 0.39A 7.2V NIMH BATTERY PACK;STANDARD ELECTRIC OUTLET - NEED TO CONFIRM EXISTING	VERATHON MEDICAL BVI3000 BLADDERSCAN	I	OF/OI
22	094399 IMU02	1	<b>IMAGING, ULTRASOUND SCANNER</b> KIOSK SYSTEM W/STAND;19" LED MONITOR;TOUCH SCREEN (WxDxH) :22x27x64 SPACE: FLOOR CM:55.9x68.6x162.6 <b>ELECT:</b> 100-240VAC; RECHARGEABLE LITHIUM-ION BATTERY - NEED TO CONFIRM EXISTING	SONOSITE, INC. L15600 X-PORTE	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, EQUIPMENT (QTY 2)  
ROOM NUMBER: 6011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
23	050708 MON03	1	<b>MONITOR, PHYSIOLOGICAL, TRANSPORT</b> BIPHASIC DEFIBRILLATOR;EXTERNAL PACING;PRINTER (WxDxH) :11x9x11 LBS: 18 SPACE: SPACE CM:27.9x22.9x27.9 KG: 8.20 <b>ELECT:</b> 120V;60HZ; - NEED TO CONFIRM EXISTING	ZOLL MEDICAL CORPORATION 43510711100124010/AC M CCT/BIPHASIC	I	OF/OI
24	079214 MON04	1	<b>MONITOR, PHYSIOLOGICAL, TRANSPORT</b> TRANSPORT MONITOR WITH MMS, TO INCLUDE: (WxDxH) :13x8x12 LBS: 13 SPACE: SPACE CM:33.0x20.3x30.5 KG: 5.90 <b>ELECT:</b> 100-240V, 50/60HZ, 70W AVERAGE, 1.2AMP <b>PLUMB:</b> NO PLUMBING REQUIRED - NEED TO CONFIRM EXISTING	PHILIPS MEDICAL SYSTEMS 866060 MX400	I	OF/OI
25	082037 ECG01	1	<b>ELECTROCARDIOGRAPH</b> WIRELESS;15IN CLR TOUCH SCREEN;12 LEADS (WxDxH) :16x13x63 LBS: 28 SPACE: FLOOR CM:40.6x33.0x160.0 KG: 12.70 <b>ELECT:</b> 120 VAC;60HZ;75W MAX - NEED TO CONFIRM EXISTING	PHILIPS MEDICAL SYSTEMS PAGEWRITER TC70 806315/860318	I	OF/OI
26	078101	2	<b>LIFTER, PATIENT, ACCESSORY, MOTOR</b> 1000LB LIFT MOTOR, 24VDC, CEILING RAIL MOUNTED (WxDxH) :14x11x8 LBS: 25 SPACE: SPACE CM:35.6x27.9x20.3 KG: 11.10 <b>ELECT:</b> 24VDC, 1.5A	PRISM MEDICAL C1000 MOTOR C1000 MOTOR	N	OF/OI

END OF ROOM





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, LINEN (QTY 2)  
ROOM NUMBER: 6012

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	096814 CRT03	2	<b>CART, LINEN</b> SS;3 FIXED SHELF;6" CASTERS (WxDxH) :62x24x70 LBS: 125 SPACE: FLOOR CM:157.5x61.0x177.8 KG: 56.70 - PATIENT LINENS	G.S. MANUFACTURING GSM-G160 -	N	OF/OI
2	006698 WST02	1	<b>WASTE RECEPTACLE</b> 44 GALLON;ROUND W/FLAT LID;GRAY CONTAINER;DOLLY (WxDxH) :24x24x37 LBS: 24 SPACE: FLOOR CM:61.0x61.0x94.0 KG: 10.90 SPECIFY COLOR: LID - FOR CLEAN GOWNS	RUBBERMAID COMM. PRODUCTS 2643-60/2645-60/2640 BRUTE	N	OF/OI
3	078147 CAB01	2	<b>CABINET, WARMING</b> MOBILE;2 COMPARTMENTS;20.6CF;GLASS DOORS;SEPARATE (WxDxH) :32x25x74 LBS: 358 SPACE: FLOOR CM:81.3x63.5x188.0 KG: 162.40 <b>ELECT:</b> 120V;50/60HZ;1PH;15A;1.8KW NEMA 5-20P;20A-125V PLUG HOSPITAL GRADE SPECIFY DOOR HINGE;RIGHT OR LEFT - PLACEHOLDER FOR FULL SIZE, SINGLE COMPARTMENT	PEDIGO PRODUCTS, INC. P-2055 P-2055	N	OF/OI
4	002532 CRT06	2	<b>CART, UTILITY</b> 3 CHROME WIRE SHELVES;5" SWIVEL CASTERS;(2) ONE (WxDxH) :36x24x38 LBS: 39 SPACE: FLOOR CM:91.4x61.0x96.5 KG: 17.70	INTERMETRO INDUSTRIES CORP. 3SPN53DC 3SP SERIES	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, NUTRITION (QTY 2)  
ROOM NUMBER: 6013

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	095453	2	<b>CART, FOOD</b>	LAKESIDE MFG. CO.	N	OF/OI
	CRT02		SS;6 TRAY CAPACITY;6"LEDGE SPACING;REMOVABLE DOOR (WxDxH) :33x22x46 LBS: 110 SPACE: FLOOR CM:83.8x55.9x116.9 KG: 49.90	SP-6373 ELITE		
			- SOILED TRAY CART			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: SOILED  
ROOM NUMBER: 6015

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	094910 COM01	1	<b>COMPACTOR, WASTE</b> 20:1 COMPACT RATION;COMMERCIAL;ELECTROHYDRAULIC (WxDxH) :43x31x90 LBS: 1210 SPACE: FLOOR CM:109.2x78.7x228.6 KG: 548.80 <b>ELECT:</b> 115V;60HZ;15 AMP;SINGLE PHASE;DEDICATED CIRCUIT - ECO-PACK	FOUNTAIN INDUSTRIES ECO-PACK 3600	N	OF/OI
2	079188 SHV01	1	<b>SHELVING, WIRE</b> 5 SHELVES;STATIONARY;CHROME;SHELF INLAY (WxDxH) :24x24x74 SPACE: FLOOR CM:61.0x61.0x188.0	INTERMETRO INDUSTRIES CORP. A2424NC/74P A2424NC	N	OF/OI
3	073358 TRK01	1	<b>TRUCK/BIN, LINEN</b> HORIZONTAL SHELVES;POLYETHYLENE;IMPACT RESISTANT (WxDxH) :48x29x64 SPACE: FLOOR CM:121.9x73.7x162.6	MEESE ORBITRON DUNNE CO. POLY TRUX 90P	N	OF/OI
4	066897	1	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0	MEDLINE INDUSTRIES, INC. MDS89668XW	N	OF/OI
5	096813	1	<b>WASTE RECEPTACLE, STEP-ON</b> 12 GAL;WHITE;FOOT PEDAL OPERATED;FIRE SAFE (WxDxH) :12x12x23 SPACE: FLOOR CM:30.5x30.5x58.4	RUBBERMAID COMM. PRODUCTS RCPST12ERBWHI DEFENDERS	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: SOILED  
ROOM NUMBER: 6015

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
6	071200	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> 9 GALLON;CLEAR SLIDE TOP;RECYCLED PLASTIC-BLACK (WxDxH) :18x12x19 SPACE: SPACE CM:45.7x30.5x48.3 SPECIFY ACCESSORIES AS SEPARATE LINE ITEMS	BECTON DICKINSON RCRA HAZARDOUS WASTE 305069	N	OF/OI
7	070664	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> CHEMOTHERAPY;PLASTIC;18 GAL;YELLOW;SLIDING LID (WxDxH) :18x13x26 LBS: 8 SPACE: FLOOR CM:45.7x33.0x66.0 KG: 3.60	KENDALL HEALTHCARE, COVIDIEN 8939 CHEMO DISPOSAL	N	OF/OI
8	089948	1	<b>WASTE RECEPTACLE, RECYCLING</b> 23 GALLON;DARK BLUE;RECYCLE SYMBOL;W/ LID (WxDxH) :20x11x35 SPACE: FLOOR CM:50.8x27.9x88.9	RUBBERMAID COMM. PRODUCTS H-1385BLU SLIMJIM 3540-07	N	OF/OI
9	079178	1	<b>WASTE RECEPTACLE</b> CONFIDENTIAL;SIDE PAPER SLOT;KEYPAD LOCK;SHIPS (WxDxH) :11x24x25 LBS: 30 SPACE: FLOOR CM:27.9x61.0x63.5 KG: 13.60	RUBBERMAID COMM. PRODUCTS 9W25 SLIM JIM	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 6018

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	058112	1	<b>REFRIGERATOR/FREEZER, UPRIGHT</b> 115V;18.2 CF;TOP FREEZER;NON-ICE;LEFT HINGE;WHITE (WxDxH) :30x32x67 LBS: 205 SPACE: FLOOR CM:76.2x81.3x170.2 KG: 93.00 <b>ELECT:</b> 115V;4.5A;60HZ;407 WATTS	SEARS COMMERCIAL SALES KENMORE 46-68972	N	OF/OI
2	087445	1	<b>ICE MAKER W/WATER DISPENSER</b> 12# STORAGE CAP;425#/24HR;LEVER;W/O FILTER (WxDxH) :17x24x33 LBS: 199 SPACE: COUNTER CM:43.2x61.0x83.8 KG: 90.30 <b>ELECT:</b> 115VAC/60HZ/1PH/11A/.8KW 8.5' CORD W/NEMA 5-15 HOSPITAL GRADE PLUG <b>HVAC:</b> 5000 BTU/HR ----- CLEARANCES: VENTILATION 6" TOP & RIGHT SIDES SERVICE 12" TOP <b>PLUMB:</b> REFER TO VENDOR TECHNICAL INFORMATION	FOLLETT CORPORATION 12CI425A-L AIR COOLED	N	OF/CI
3	080483	1	<b>OVEN, MICROWAVE</b> 1000W;S.S. INTERIOR/EXTERIOR (WxDxH) :21x16x13 LBS: 37 SPACE: COUNTER CM:53.3x40.6x33.0 KG: 16.80 <b>ELECT:</b> 120V;14A/11;1.6KW;15AMP NEMA 15-R RECEPTACLE	SHARP ELECTRONICS CORP. R21LCF R21LCF	N	OF/OI



A division of Ross & Baruzzini

712A

**MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT**

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 6018

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
4	062450 COF01	1	<b>COFFEE BREWER</b> PLUMBED;1 LOWER BREWER/2 UPPER WARMERS;200 OZ CAP. (WxDxH) :9x22x19 LBS: 30 SPACE: COUNTER CM:22.9x55.9x48.3 KG: 13.60 <b>ELECT:</b> 120V/15AMPS;TANK HEATER 1425W;TOTAL WATTS 1800 CORD ATTACHED;REQ.2-WIRES PLUS GROUND SERVICE RATED 120V,SINGLE PHASE,60HZ. <b>PLUMB:</b> 20-90 PSI (138-621 KPA).MACHINES SUPPLIED W/ 1/4" MALE FLARE FITTING - TO INCLUDE HOT WATER TAP	BUNN-O-MATIC CORP. AXIOM-15-3-0000/ACC 38700.0000	L	OF/VI
5	051789 TST01	1	<b>TOASTER, COMMERCIAL</b> 2-SLOT;CHROME-PLATED CONSTRUCTION;1 3/8"WIDE SLOTS (WxDxH) :14x9x9 LBS: 5 SPACE: COUNTER CM:35.6x22.9x22.9 KG: 2.30 <b>ELECT:</b> 120 VOLTS/8 AMPS/925 WATTS/PHASE 1	WARING PRODUCTS WCT702 -	N	OF/OI
6	094915 DSP02	1	<b>DISPENSER, FOOD SERVICE, BEVERAGE</b> SOFT DRINK;4 POST MIXING DISPENSING VALVES;LEG KIT (WxDxH) :11x26x28 LBS: 145 SPACE: COUNTER CM:27.9x66.0x71.1 KG: 65.80 <b>ELECT:</b> 120V;60HZ;7 AMPS <b>PLUMB:</b> SS LABELED INLET SYRUP/WATER;3/8"BARB - COORDINATION REQUIRED FOR SYRUP, CO2, ETC UNDER COUNTER	LANCER CORP. CED 500	L	OF/OI
7	061957 GLV02	1	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 SPACE: WALL CM:45.7x10.2x22.9	CLINTON INDUSTRIES G-1030 -	N	OF/CI



A division of **Ross&Baruzzini**

712A

**MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT**

-

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 6018

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
8	006677	1	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 LBS: 43 SPACE: FLOOR CM:30.5x53.3x88.9 KG: 19.50 SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI

**END OF ROOM**



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, CART  
ROOM NUMBER: 6022

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - IV CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	N	OF/OI
2	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - RT CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	N	OF/OI
3	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - FOR LINEN	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	N	OF/OI
4	092750 MON05	1	<b>MONITOR, VITAL SIGNS</b> NIBP;SPO2;TEMPERATURE;RECORDER;ADULT/PEDIATRIC/ (WxDxH) :25x25x50 LBS: 12 SPACE: FLOOR CM:63.5x63.5x127.0 KG: 5.40 <b>ELECT:</b> 120 VOLT/60 HZ BATTERY CHARGER 10.8-11.1V LITHIUM ION BATTERY - FOR AMBULATION	PHILIPS MEDICAL SYSTEMS 863283/989803176601 SURE SIGNS VS4	N	OF/OI

END OF ROOM





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: MEDICATION  
ROOM NUMBER: 6023

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	072239	2	<b>DISPENSER, MEDICATION, AUTOMATED</b> 14.75"LCD MONITOR;KEYBOARD;TOUCHPAD;BIO-ID SCANNER (WxDxH) :23x27x55 LBS: 166 SPACE: FLOOR CM:58.4x68.6x139.7 KG: 75.30 <b>ELECT:</b> 120V;60HZ;3 AMPS MAX EMERGENCY POWER REQUIRED <b>HVAC:</b> 409 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET - NEED TO UPDATE TO ES SERIES	BD CAREFUSION PYXIS MEDSTATION ES 6 DRAWER MAIN	N	OF/VI
2	072241	2	<b>DISPENSER, MEDICATION, AUTOMATED</b> SLIDE BRACKETS W/DRAWER RELEASE;PLASTIC TOP (WxDxH) :23x47x47 LBS: 133 SPACE: FLOOR CM:58.4x119.4x119.4 KG: 60.30 <b>ELECT:</b> 120V;60HZ;3 AMPS MAX;EMERGENCY POWER REQUIRED <b>HVAC:</b> 222 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET	BD CAREFUSION PYXIS 7 DRAWER AUXILIARY MEDSTATION ES	N	OF/OI
3	072243	2	<b>DISPENSER, MEDICATION, AUTOMATED</b> SINGLE COLUMN;1 STORAGE SHELF PER DOOR;INTERIOR (WxDxH) :31x28x80 LBS: 314 SPACE: FLOOR CM:78.7x71.1x203.2 KG: 142.40 <b>ELECT:</b> 120V;60HZ;3 AMPS;EMERGENCY POWER REQUIRED <b>HVAC:</b> 222 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET	BD CAREFUSION PYXIS FOUR DOOR AUXILIARY MEDSTATION ES	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: MEDICATION  
ROOM NUMBER: 6023

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
4	079417 REF01	2	<b>REFRIGERATOR, UPRIGHT, MEDICAL GRADE</b> 115V;SINGLE DOOR;19.7 CF;SS;LED DIG. TEMP;LOCKING (WxDxH) :30x31x80 LBS: 395 SPACE: FLOOR CM:76.2x78.7x203.2 KG: 179.20 <b>ELECT:</b> 115V,60HZ,8.7A <b>HVAC:</b> 950 BTU/HR MAX HEAT REJECTION	FOLLETT CORPORATION REF20-LB-R000 REF20-LB-R000	N	OF/OI
5	053907 MED01	2	<b>DISPENSER, MEDICATION, ACCESSORY</b> MED DISPENSING UNIT REFRIGERATOR CONTROL (WxDxH) :3x6x7 LBS: 4 SPACE: SPACE CM:7.6x15.2x17.8 KG: 1.80 <b>ELECT:</b> POWERED FROM MAIN DISPENSER	BD CAREFUSION PYXIS REMOTE MANAGER MSRM	N	OF/VI
6	506535 OSS01	1	<b>OPEN SUPPLY SYSTEM</b>  - PAREX SUPPLY SYSTEM, WEIGHT BASED -PROVIDE BACKING ON ALL WALLS WITH NO OUTLETS EXTENDING BEYOND 12" AFF. REQUIRES POWER AND DATA AT CONTROL LOCATION.	ZZZ - GENERIC DESCRIPTIONS	N	OF/VI
7	077862 BIN01	8	<b>BIN, ACCESSORY, PANEL, LOUVERED</b> LOUVERED PANEL;36";STONE (WxDxH) :36x1x19 SPACE: WALL CM:91.4x2.5x48.3	PAR EXCELLENCE SYSTEMS, INC. 30-636	N	OF/CI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: MEDICATION  
ROOM NUMBER: 6023

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
8	093412 MVD01	1	<b>MONITOR, VIDEO</b> MXU0299: DISPLAY FLAT PANEL NON-TOUCH PIIC SPACE: SPACE	PHILIPS MEDICAL SYSTEMS MXU0299 866390	N	OF/OI
- TO INCLUDE SPEAKERS						
9	055704 EWS01	1	<b>ENCLOSURE, WALL, SHARPS</b> 2 OR 5 QT WALL ENCLOSURE (WxDxH) :13x6x12 CM:33.0x15.2x30.5 SPACE: WALL	KENDALL HEALTHCARE, COVIDIEN 85161H IN-ROOM	L	OF/CI
10	093535 DSL02	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> 2GAL;BLACK;WHITE HINGED LID W/PORT;LEAK RESISTANT (WxDxH) :12x10x12 CM:30.5x25.4x30.5 SPACE: WALL	KENDALL HEALTHCARE, COVIDIEN 8602RC/8963 8602RC/8963	N	OF/OI
11	061957 GLV02	1	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 CM:45.7x10.2x22.9 SPACE: WALL	CLINTON INDUSTRIES G-1030 -	N	OF/CI
12	006677	1	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 CM:30.5x53.3x88.9 LBS: 43 KG: 19.50 SPACE: FLOOR SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: MEDICATION  
ROOM NUMBER: 6023

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
13	047636	2	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60 - PLACEHOLDER FOR RED AND BLACK BINS. NEED VENDOR AND MODEL.	RUBBERMAID COMM. PRODUCTS 2544 40 QT	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: DICTATION  
ROOM NUMBER: 6027

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	096814 CRT03	1	<b>CART, LINEN</b> SS;3 FIXED SHELF;6" CASTERS (WxDxH) :62x24x70 LBS: 125 SPACE: FLOOR CM:157.5x61.0x177.8 KG: 56.70 - GOWN	G.S. MANUFACTURING GSM-G160 -	N	OF/OI
2	094909 CRT07	1	<b>CART, UTILITY</b> MULTIPURPOSE CART;800LB CAPACITY;5"WHEELS (WxDxH) :21x31x42 LBS: 44 SPACE: FLOOR CM:53.3x78.7x106.7 KG: 20.00	VERSACART EZTOTE 390 101-390	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: CLEAN SUPPLY  
ROOM NUMBER: 6028

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	506535 OSS01	1	<b>OPEN SUPPLY SYSTEM</b>	ZZZ - GENERIC DESCRIPTIONS	N	OF/VI
			- PAREX SUPPLY SYSTEM, WEIGHT BASED -PROVIDE BACKING ON ALL WALLS WITH NO OUTLETS EXTENDING BEYOND 12" AFF. REQUIRES POWER AND DATA AT CONTROL LOCATION.			
2	077862 BIN01	30	<b>BIN, ACCESSORY, PANEL, LOUVERED</b> LOUVERED PANEL;36";STONE (WxDxH) :36x1x19 CM:91.4x2.5x48.3	PAR EXCELLENCE SYSTEMS, INC. 30-636	N	OF/CI
			SPACE: WALL			
3	084658 CRT01	3	<b>CART, CYLINDER, MEDICAL GAS</b> 6 CAPACITY; E/D SIZE;W/HEAVY DUTY 5" CASTERS (WxDxH) :11x16x39 CM:27.9x40.6x99.1	W.T. FARLEY, INC. CR-DC06E-G -	N	OF/OI
			LBS: 22 KG: 10.00			
			SPACE: FLOOR			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, CORRIDOR  
ROOM NUMBER: 6033

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	002532 CRT06	1	<b>CART, UTILITY</b> 3 CHROME WIRE SHELVES;5" SWIVEL CASTERS;(2) ONE (WxDxH) :36x24x38 LBS: 39 SPACE: FLOOR CM:91.4x61.0x96.5 KG: 17.70	INTERMETRO INDUSTRIES CORP. 3SPN53DC 3SP SERIES	N	OF/OI
2	096814 CRT03	1	<b>CART, LINEN</b> SS;3 FIXED SHELF;6" CASTERS (WxDxH) :62x24x70 LBS: 125 SPACE: FLOOR CM:157.5x61.0x177.8 KG: 56.70 - GOWN	G.S. MANUFACTURING GSM-G160 -	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, EQUIPMENT  
ROOM NUMBER: 6035

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	089251 CRT05	1	<b>CART, RESUSCITATION, CARDIAC</b> 4 DRAWER;2 3",1 6",1 9";ADJ DEFIB TRAY;IV POLE (WxDxH) :44x24x52 CM:111.8x61.0x132.1 <b>ELECT:</b> 120V FOR POWER STRIP  -DIMENSIONS LISTED DO NOT INCLUDE SIDE ACCESSORIES, IV POLE HEIGHT, OR DEFIBRILLATOR SHELF HEIGHT	INTERMETRO INDUSTRIES CORP. LECCRP3 LIFELINE	N	OF/OI
2	087848 DEF01	1	<b>DEFIBRILLATOR</b> GUIDELINES 2015 - COMPATIBLE, AED W/MANUAL OVERRIDE (WxDxH) :11x13x8 CM:27.9x33.0x20.3 <b>ELECT:</b> 120VAC;60HZ;2AMPS;SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER	ZOLL MEDICAL CORPORATION 30120000001110012 R SERIES ALS	E	OF/OI
3	092325	1	<b>ASPIRATOR</b> VACUUM;5 SETTINGS;30LPM;BUILT IN AC/DC POWER;LED (WxDxH) :13x7x13 CM:33.0x17.8x33.0 <b>ELECT:</b> 100-240VAC;50/60HZ;RECHARGEABLE 12VDC BATTERY	LAERDAL MEDICAL CORP. 78002001/ACC LSU W/CANISTER	N	OF/OI
4	079188 SHV01	1	<b>SHELVING, WIRE</b> 5 SHELVES;STATIONARY;CHROME;SHELF INLAY (WxDxH) :24x24x74 CM:61.0x61.0x188.0	INTERMETRO INDUSTRIES CORP. A2424NC/74P A2424NC	N	OF/OI

END OF ROOM





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: SERVICE ELEVATOR LOBBY  
ROOM NUMBER: 6084

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	063113	3	<b>STRETCHER</b>	STRYKER CORP/MEDICAL DIV	I	OF/OI
	STR01		700LB CAPACITY;3"MATTRESS;ZOOM DRIVE;30" WIDE; (WxDxH) :38x85x34 SPACE: FLOOR CM:96.5x215.9x86.4	PRIME SERIES 1125 ZOOM		
			- FOR SPACE AND UTILITY PLANNING ONLY			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 6702

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	006264	1	<b>NOTE:</b>	ZZZ - NOTE: (U.S.)	I	OF/OI
			- 18 REGULAR ROOMS, 1 ISOLATION, 1 POSITIVE PRESSURE			
2	086193 BED01	20	<b>BED, MED-SURG</b> 500# CAP;IBED UPGRADEABLE;INBED SCALE;ISOFLEX (WxDxH) :43x93x30 SPACE: FLOOR CM:109.2x236.2x76.2	STRYKER CORP/MEDICAL DIV S3 PX2-3005 S3	E	OF/OI
			- *ASSUME RELOCATED FROM RICHARDS (R1, R3, R4, R5). CONFIRMATION NEEDED. CONFIRM O2 TANK HOLDER QUANTITIES			
3	053630	20	<b>BED ACCESSORIES</b> FULLY DYNAMIC;NON-PWR;PRESSURE MANAGEMENT SURFACE (WxDxH) :35x84x7 LBS: 50 SPACE: SPACE CM:88.9x213.4x17.8 KG: 22.70	B.G. INDUSTRIES ACCUMAX QUANTUM A3-117610-384	E	OF/OI
			<b>ELECT:</b> 110 VOLT/PHASE 1 - *ASSUME RELOCATED FROM RICHARDS (R1, R3, R4, R5). CONFIRMATION NEEDED.			
4	096840 LFT05	20	<b>LIFTER, PATIENT, ACCESSORY, RAIL</b> 10'STRAIGHT;FOR USE WITH C1000 MOTOR SPACE: CEILING	PRISM MEDICAL 10' STRAIGHT	N	OF/VI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 6702

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
5	086984 MON06	20	<b>MONITOR, VITAL SIGNS</b> 8.4"LCD SCREEN;NBP;SPO2;TYMPANIC TEMP;WIRELESS (WxDxH) :11x6x9 LBS: 7 SPACE: WALL CM:27.9x15.2x22.9 KG: 3.10 <b>ELECT:</b> 120 VOLT/60 HZ BATTERY CHARGER LITHIUM ION BATTERY 10.8-11.1V DATA OUTPUT:HL7 FORMAT, VIA ETHERNET PORT SERIAL DATA	PHILIPS MEDICAL SYSTEMS 863283 SURESIGNS VS4	N	OF/OI
6	093331 WCH01	20	<b>WALL CHANNEL</b> WALL CHANNEL;SEISMIC;19"  SPACE: WALL	PHILIPS MEDICAL SYSTEMS MXU175	N	OF/CI
7	077835 MNT01	20	<b>MONITOR ACCESSORY, MOUNTING ASSEMBLY</b> VARIABLE HEIGHT MOUNT;TILT/SWIVEL ADJUSTMENT;  SPACE: WALL	GCX CORPORATION WA-0011-27 VHM-25	N	OF/OI
8	028803	20	<b>STAND, MAYO</b> MOBILE;CHROME;HAND HEIGHT ADJUSTABLE 30-40 IN (WxDxH) :13x20x40 LBS: 22 SPACE: FLOOR CM:33.0x50.8x101.6 KG: 10.00 - PLACEHOLDER. SPECS TO BE DETERMINED. MAY NOT NEED 1 PER ROOM.	PEDIGO PRODUCTS, INC. P-65 MOBILE	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 6702

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
9	067194	20	<b>I.V. POLE</b> STAINLESS STEEL;5 LEG;3 IN CASTERS;4 HOOK TOP (WxDxH) :23x6x35 LBS: 32 SPACE: FLOOR CM:58.4x15.2x88.9 KG: 14.50 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	PRYOR PRODUCTS 135	N	OF/OI
10	088643	20	<b>I.V. INFUSION PUMP</b> GENERAL INFUSION PUMP;SINGLE CHANNEL;POLE CLAMP; (WxDxH) :8x8x6 LBS: 10 SPACE: SPACE CM:20.3x20.3x15.2 KG: 4.50 <b>ELECT:</b> 120V;50/60HZ;35W;1A;1 RECHARGEABLE BATTERY LEAD-ACID 6V;POWER CORD:HOSPITAL-GRADE AC CORD 10 FT LONG W/TRANSPARENT PLUG AND RETAINER PLATE - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	HOSPIRA, INC. 30010 PLUM 360/MEDNET	N	OF/OI
11	096809	20	<b>PUMP, COMPRESSION</b> DVT;LCD SCREEN;ONE BUTTON START;CARRY HANDLE (WxDxH) :9x8x10 LBS: 9 SPACE: SPACE CM:22.9x20.3x25.4 KG: 4.10 <b>ELECT:</b> 120V/60HZ - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	ARJOHUNTLEIGH ACS900 FLOWTRON	L	OF/OI
12	066897	20	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	MEDLINE INDUSTRIES, INC. MDS89668XW	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 6702

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
13	087247	20	<b>WALKER, FOLDING, ADULT</b> SIDE WALKER;ALUMINUM;USER WT CAP 300LBS;HT ADJUST (WxDxH) :19x16x36 LBS: 4 SPACE: FLOOR CM:48.3x40.6x91.4 KG: 1.60 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	PERFORMANCE HEALTHCARE 081561752 HEMI DAYS INVACARE	N	OF/OI
14	063659	20	<b>TABLE, OVERBED</b> SINGLE TOP;W/VANITY;LOW PROFILE U-SHAPED BASE (WxDxH) :21x32x45 LBS: 53 SPACE: FLOOR CM:53.3x81.3x114.3 KG: 24.10 SPECIFY FINISH SPECIFY FINISH: BASE SPECIFY FINISH: EDGE	STRYKER CORP/MEDICAL DIV 3150-000-200 TRU-FIT	E	OF/OI
15	080275	40	<b>REGULATOR, SUCTION, CONTINUOUS &amp; INTERMITTENT</b> 0-200MMHG W/SELECTOR VALVE (OFF,INT,CONT);CANISTER (WxDxH) :4x6x4 LBS: 1 SPACE: WALL CM:10.2x15.2x10.2 KG: 0.50	BOEHRINGER LABORATORIES, INC. 3804/1496/2469 PLATINUM SERIES	E	OF/OI
16	055048	20	<b>FLOWMETER, OXYGEN</b> W/DISS ADAPTOR  SPACE: SPACE  CONFIRM MEDICAL GAS CONNECTION	AMVEX, DIVISION OF OHIO MEDICAL FM15ODH -	E	OF/OI
17	062114 DSP01	20	<b>DISPENSER</b> EMEBAG DISPENSER (WxDxH) :7x7x7 SPACE: WALL CM:17.8x17.8x17.8	CENTURION MEDICAL PRODUCTS EMED200 EMEBAG	N	OF/CI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 6702

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
18	091196 DSP03	20	<b>DISPENSER, MASK</b> PLEXIGLAS DISPENSER FOR SURGICAL FACE MASK BOXES (WxDxH) :9x6x9 LBS: 2 SPACE: WALL CM:22.9x15.2x22.9 KG: 0.90	MEDLINE INDUSTRIES, INC. NONFM122 NONFM122	N	OF/OI
19	087029 DSL01	20	<b>DISPOSAL CONTAINER, SHARPS W/BACKET</b> LOCKING WALL CABINET FOR SHARPS CONTAINERS;2 GAL (WxDxH) :15x6x14 SPACE: WALL CM:38.1x15.2x35.6	STERICYCLE OC-02-2004 BIO SYSTEMS	L	OF/CI
20	061957 GLV02	20	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 SPACE: WALL CM:45.7x10.2x22.9	CLINTON INDUSTRIES G-1030 -	N	OF/CI
21	058768 HMP01	20	<b>HAMPER, LINEN</b> W/LID;FOOT PEDAL;STEEL FRAME;POWDER COAT FINISH (WxDxH) :21x22x33 SPACE: FLOOR CM:53.3x55.9x83.8 - INSTALLED IN CASEWORK	CENTURION MEDICAL PRODUCTS CX302 MEDIUM	N	OF/OI
22	078043 HMP02	20	<b>HAMPER, LINEN</b> W/LID;FOOT PEDAL OPERATED;TUBULAR CHROME FRAME; (WxDxH) :19x20x26 SPACE: FLOOR CM:48.3x50.8x66.0 - FOR YELLOW BAG CHEMO. INSTALLED IN CASEWORK	CENTURION MEDICAL PRODUCTS CENTURION CX502	N	OF/OI
23	047636	20	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60	RUBBERMAID COMM. PRODUCTS 2544 40 QT	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 6702

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
24	085429 RLS01	40	<b>RAIL SYSTEM</b> 7"2" VERTICAL RAIL SYSTEM;SURFACE MOUNT;1 TIER SPACE: WALL	MODULAR SERVICES COMPANY 530-4522	N	CF/CI
			- WITH BASKET, SHELF, AND OTHER ACCESSORIES TO BE DETERMINED.			
25	504469	20	<b>CLOCK</b>	ZZZ - GENERIC DESCRIPTIONS	N	OF/OI
			- SPECIFIED BY OTHERS			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-1 ONCOLOGY-SURGICAL  
ROOM NAME: TOILET, PATIENT (QTY 20)  
ROOM NUMBER: 6702A

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	047636	20	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60	RUBBERMAID COMM. PRODUCTS 2544 40 QT	N	OF/OI

END OF ROOM  
END OF DEPARTMENT





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: STAFF LOUNGE  
ROOM NUMBER: 6049

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	058112	2	<b>REFRIGERATOR/FREEZER, UPRIGHT</b> REF02 115V;18.2 CF;TOP FREEZER;NON-ICE;LEFT HINGE;WHITE (WxDxH) :30x32x67 LBS: 205 SPACE: FLOOR CM:76.2x81.3x170.2 KG: 93.00 ELECT: 115V;4.5A;60HZ;407 WATTS	SEARS COMMERCIAL SALES KENMORE 46-68972	N	OF/OI
2	080483	2	<b>OVEN, MICROWAVE</b> OVE01 1000W;S.S. INTERIOR/EXTERIOR (WxDxH) :21x16x13 LBS: 37 SPACE: COUNTER CM:53.3x40.6x33.0 KG: 16.80 ELECT: 120V;14A/11;1.6KW;15AMP NEMA 15-R RECEPTACLE	SHARP ELECTRONICS CORP. R21LCF R21LCF	N	OF/OI
3	090442	1	<b>COFFEE BREWER</b> COF02 SINGLE-CUP BREWER FOR COFFEE, TEA, HOT CHOCOLATE (WxDxH) :11x14x14 LBS: 18 SPACE: COUNTER CM:27.9x35.6x35.6 KG: 8.20 ELECT: 120VAC, 1400 WATTS - TO INCLUDE HOT WATER TAP	KEURIG, INC. COMMERCIAL BREWING K150	L	OF/VI
4	051789	1	<b>TOASTER, COMMERCIAL</b> TST01 2-SLOT;CHROME-PLATED CONSTRUCTION;1 3/8"WIDE SLOTS (WxDxH) :14x9x9 LBS: 5 SPACE: COUNTER CM:35.6x22.9x22.9 KG: 2.30 ELECT: 120 VOLTS/8 AMPS/925 WATTS/PHASE 1	WARING PRODUCTS WCT702 -	N	OF/OI
5	006677	2	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 LBS: 43 SPACE: FLOOR CM:30.5x53.3x88.9 KG: 19.50 SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI



A division of **Ross&Baruzzini**

712A

**MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT**

-

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: STAFF LOUNGE  
ROOM NUMBER: 6049

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
6	079040	1	<b>CHAIR, MASSAGE</b> 6 MOTOR;2 AIR PUMPS;FULL BODY;RECLINER;LCD SCREEN (WxDxH) :31x53x49 LBS: 220 SPACE: FLOOR CM:78.7x134.6x124.5 KG: 99.80 <b>ELECT:</b> AC 120V;60HZ;20-230W	BEAUTYHEALTH BC-07DH	N	OF/OI

**END OF ROOM**



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: CORRIDOR  
ROOM NUMBER: 6051

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	093412 MVD01	4	<b>MONITOR, VIDEO</b> MXU0299: DISPLAY FLAT PANEL NON-TOUCH PIIC SPACE: SPACE	PHILIPS MEDICAL SYSTEMS MXU0299 866390	N	OF/OI
2	093331 WCH01	4	<b>WALL CHANNEL</b> WALL CHANNEL;SEISMIC;19" SPACE: WALL	PHILIPS MEDICAL SYSTEMS MXU175	N	OF/CI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT CARE STATION (QTY 3)  
ROOM NUMBER: 6052

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	087769 MON01	1	<b>MONITOR, CENTRAL STATION</b> PIIC IX SURVEILLANCE CENTRAL MONITORING;19" TOUCH (WxDxH) :48x16x16 SPACE: COUNTER CM:121.9x40.6x40.6 <b>ELECT:</b> CPU: 115 VAC;60 HZ;5.2 AMPS LCD DISPLAY: 115 VAC;60 HZ;48W TYP. UPS: 115 VAC;60 HZ;5.6 AMPS PRINTER: 117 VAC;60 HZ;5.3 AMPS <b>HVAC:</b> MUST MAINTAIN TEMP AT 62-82 DEGREES F AND NON-CONDENSING RELATIVE HUMIDITY AT 30-60% CPU: 215 BTU/HR UPS: 437 BTU/HR (MAX) DATA CONNECTION REQUIRED, SEE TECH SHEET	PHILIPS MEDICAL SYSTEMS INTELLIVUE INFO CENTER IX	N	OF/OI
2	504528 MON17	1	<b>ALLOWANCE, MONITORING, TELEMETRY</b>	ZZZ - MONITORING	N	OF/OI
- ALLOWANCE FOR TELEMETRY INFRASTRUCTURE						
3	086723 ANA01	2	<b>LAB ANALYZER, GLUCOSE</b> BLOOD GLUCOSE & KETONE MONITORING;BARCODE READER (WxDxH) :3x2x8 LBS: 1 SPACE: SPACE CM:7.6x5.1x20.3 KG: 0.50 <b>ELECT:</b> 2 AA BATTERIES (ALKALINE, LITHIUM, OR NICAD)	ABBOTT DIAGNOSTICS 71411-70 PRECISION PRO	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL

ROOM NAME: MEDICATION

ROOM NUMBER: 6059

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	072239	1	<b>DISPENSER, MEDICATION, AUTOMATED</b>	BD CAREFUSION PYXIS	N	OF/VI
	MED02		14.75"LCD MONITOR;KEYBOARD;TOUCHPAD;BIO-ID SCANNER (WxDxH) :23x27x55 LBS: 166 SPACE: FLOOR CM:58.4x68.6x139.7 KG: 75.30	MEDSTATION ES 6 DRAWER MAIN		
			<b>ELECT:</b> 120V;60HZ;3 AMPS MAX EMERGENCY POWER REQUIRED			
			<b>HVAC:</b> 409 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET - NEED TO UPDATE TO ES SERIES			
2	072241	1	<b>DISPENSER, MEDICATION, AUTOMATED</b>	BD CAREFUSION PYXIS	N	OF/OI
	MED05		SLIDE BRACKETS W/DRAWER RELEASE;PLASTIC TOP (WxDxH) :23x47x47 LBS: 133 SPACE: FLOOR CM:58.4x119.4x119.4 KG: 60.30	7 DRAWER AUXILIARY MEDSTATION ES		
			<b>ELECT:</b> 120V;60HZ;3 AMPS MAX;EMERGENCY POWER REQUIRED			
			<b>HVAC:</b> 222 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET			
3	072243	1	<b>DISPENSER, MEDICATION, AUTOMATED</b>	BD CAREFUSION PYXIS	N	OF/OI
	MED03		SINGLE COLUMN;1 STORAGE SHELF PER DOOR;INTERIOR (WxDxH) :31x28x80 LBS: 314 SPACE: FLOOR CM:78.7x71.1x203.2 KG: 142.40	FOUR DOOR AUXILIARY MEDSTATION ES		
			<b>ELECT:</b> 120V;60HZ;3 AMPS;EMERGENCY POWER REQUIRED			
			<b>HVAC:</b> 222 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET			



A division of Ross & Baruzzini

712A

**MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT**

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: MEDICATION  
ROOM NUMBER: 6059

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
4	079417 REF01	1	<b>REFRIGERATOR, UPRIGHT, MEDICAL GRADE</b> 115V;SINGLE DOOR;19.7 CF;SS;LED DIG. TEMP;LOCKING (WxDxH) :30x31x80 LBS: 395 SPACE: FLOOR CM:76.2x78.7x203.2 KG: 179.20 <b>ELECT:</b> 115V,60HZ,8.7A <b>HVAC:</b> 950 BTU/HR MAX HEAT REJECTION	FOLLETT CORPORATION REF20-LB-R000 REF20-LB-R000	N	OF/OI
5	053907 MED01	1	<b>DISPENSER, MEDICATION, ACCESSORY</b> MED DISPENSING UNIT REFRIGERATOR CONTROL (WxDxH) :3x6x7 LBS: 4 SPACE: SPACE CM:7.6x15.2x17.8 KG: 1.80 <b>ELECT:</b> POWERED FROM MAIN DISPENSER	BD CAREFUSION PYXIS REMOTE MANAGER MSRM	N	OF/VI
6	506535 OSS01	1	<b>OPEN SUPPLY SYSTEM</b>  - PAREX SUPPLY SYSTEM, WEIGHT BASED -PROVIDE BACKING ON ALL WALLS WITH NO OUTLETS EXTENDING BEYOND 12" AFF. REQUIRES POWER AND DATA AT CONTROL LOCATION.	ZZZ - GENERIC DESCRIPTIONS	N	OF/VI
7	077862 BIN01	2	<b>BIN, ACCESSORY, PANEL, LOUVERED</b> LOUVERED PANEL;36";STONE (WxDxH) :36x1x19 SPACE: WALL CM:91.4x2.5x48.3	PAR EXCELLENCE SYSTEMS, INC. 30-636	N	OF/CI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: MEDICATION  
ROOM NUMBER: 6059

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
8	093412 MVD01	1	<b>MONITOR, VIDEO</b> MXU0299: DISPLAY FLAT PANEL NON-TOUCH PIIC SPACE: SPACE	PHILIPS MEDICAL SYSTEMS MXU0299 866390	N	OF/OI
- TO INCLUDE SPEAKERS						
9	055704 EWS01	1	<b>ENCLOSURE, WALL, SHARPS</b> 2 OR 5 QT WALL ENCLOSURE (WxDxH) :13x6x12 CM:33.0x15.2x30.5 SPACE: WALL	KENDALL HEALTHCARE, COVIDIEN 85161H IN-ROOM	L	OF/CI
10	093535 DSL02	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> 2GAL;BLACK;WHITE HINGED LID W/PORT;LEAK RESISTANT (WxDxH) :12x10x12 CM:30.5x25.4x30.5 SPACE: WALL	KENDALL HEALTHCARE, COVIDIEN 8602RC/8963 8602RC/8963	N	OF/OI
11	061957 GLV02	1	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 CM:45.7x10.2x22.9 SPACE: WALL	CLINTON INDUSTRIES G-1030 -	N	OF/CI
12	006677	1	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 CM:30.5x53.3x88.9 LBS: 43 KG: 19.50 SPACE: FLOOR SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: MEDICATION  
ROOM NUMBER: 6059

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
13	047636	2	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60 - PLACEHOLDER FOR RED AND BLACK BINS. NEED VENDOR AND MODEL.	RUBBERMAID COMM. PRODUCTS 2544 40 QT	N	OF/OI

END OF ROOM





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: CLEAN SUPPLY  
ROOM NUMBER: 6063

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	506535 OSS01	1	<b>OPEN SUPPLY SYSTEM</b>	ZZZ - GENERIC DESCRIPTIONS	N	OF/VI
			- PAREX SUPPLY SYSTEM, WEIGHT BASED -PROVIDE BACKING ON ALL WALLS WITH NO OUTLETS EXTENDING BEYOND 12" AFF. REQUIRES POWER AND DATA AT CONTROL LOCATION.			
2	077862 BIN01	30	<b>BIN, ACCESSORY, PANEL, LOUVERED</b> LOUVERED PANEL;36";STONE (WxDxH) :36x1x19 CM:91.4x2.5x48.3	PAR EXCELLENCE SYSTEMS, INC. 30-636	N	OF/CI
			SPACE: WALL			
3	084658 CRT01	3	<b>CART, CYLINDER, MEDICAL GAS</b> 6 CAPACITY; E/D SIZE;W/HEAVY DUTY 5" CASTERS (WxDxH) :11x16x39 CM:27.9x40.6x99.1	W.T. FARLEY, INC. CR-DC06E-G -	N	OF/OI
			LBS: 22 KG: 10.00			
			SPACE: FLOOR			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: SOILED HOLDING  
ROOM NUMBER: 6064

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	094910 COM01	1	<b>COMPACTOR, WASTE</b> 20:1 COMPACT RATION;COMMERCIAL;ELECTROHYDRAULIC (WxDxH) :43x31x90 LBS: 1210 SPACE: FLOOR CM:109.2x78.7x228.6 KG: 548.80 <b>ELECT:</b> 115V;60HZ;15 AMP;SINGLE PHASE;DEDICATED CIRCUIT - ECO-PACK	FOUNTAIN INDUSTRIES ECO-PACK 3600	N	OF/OI
2	079188 SHV01	1	<b>SHELVING, WIRE</b> 5 SHELVES;STATIONARY;CHROME;SHELF INLAY (WxDxH) :24x24x74 SPACE: FLOOR CM:61.0x61.0x188.0	INTERMETRO INDUSTRIES CORP. A2424NC/74P A2424NC	N	OF/OI
3	073358 TRK01	1	<b>TRUCK/BIN, LINEN</b> HORIZONTAL SHELVES;POLYETHYLENE;IMPACT RESISTANT (WxDxH) :48x29x64 SPACE: FLOOR CM:121.9x73.7x162.6	MEESE ORBITRON DUNNE CO. POLY TRUX 90P	N	OF/OI
4	066897	1	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0	MEDLINE INDUSTRIES, INC. MDS89668XW	N	OF/OI
5	096813	1	<b>WASTE RECEPTACLE, STEP-ON</b> 12 GAL;WHITE;FOOT PEDAL OPERATED;FIRE SAFE (WxDxH) :12x12x23 SPACE: FLOOR CM:30.5x30.5x58.4	RUBBERMAID COMM. PRODUCTS RCPST12ERBWHI DEFENDERS	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: SOILED HOLDING  
ROOM NUMBER: 6064

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
6	071200	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> 9 GALLON;CLEAR SLIDE TOP;RECYCLED PLASTIC-BLACK (WxDxH) :18x12x19 SPACE: SPACE CM:45.7x30.5x48.3 SPECIFY ACCESSORIES AS SEPARATE LINE ITEMS	BECTON DICKINSON RCRA HAZARDOUS WASTE 305069	N	OF/OI
7	070664	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> CHEMOTHERAPY;PLASTIC;18 GAL;YELLOW;SLIDING LID (WxDxH) :18x13x26 LBS: 8 SPACE: FLOOR CM:45.7x33.0x66.0 KG: 3.60	KENDALL HEALTHCARE, COVIDIEN 8939 CHEMO DISPOSAL	N	OF/OI
8	089948	1	<b>WASTE RECEPTACLE, RECYCLING</b> 23 GALLON;DARK BLUE;RECYCLE SYMBOL;W/ LID (WxDxH) :20x11x35 SPACE: FLOOR CM:50.8x27.9x88.9	RUBBERMAID COMM. PRODUCTS H-1385BLU SLIMJIM 3540-07	N	OF/OI
9	079178	1	<b>WASTE RECEPTACLE</b> CONFIDENTIAL;SIDE PAPER SLOT;KEYPAD LOCK;SHIPS (WxDxH) :11x24x25 LBS: 30 SPACE: FLOOR CM:27.9x61.0x63.5 KG: 13.60	RUBBERMAID COMM. PRODUCTS 9W25 SLIM JIM	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, CART (QTY 2)  
ROOM NUMBER: 6065

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	092815 CRT04	2	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - IV CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	N	OF/OI
2	092815 CRT04	2	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - RT CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	N	OF/OI
3	092815 CRT04	2	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - FOR LINEN	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	N	OF/OI
4	092750 MON05	2	<b>MONITOR, VITAL SIGNS</b> NIBP;SPO2;TEMPERATURE;RECORDER;ADULT/PEDIATRIC/ (WxDxH) :25x25x50 LBS: 12 SPACE: FLOOR CM:63.5x63.5x127.0 KG: 5.40 <b>ELECT:</b> 120 VOLT/60 HZ BATTERY CHARGER 10.8-11.1V LITHIUM ION BATTERY - FOR AMBULATION	PHILIPS MEDICAL SYSTEMS 863283/989803176601 SURE SIGNS VS4	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, CART (QTY 2)  
ROOM NUMBER: 6065

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
5	089251 CRT05	2	<b>CART, RESUSCITATION, CARDIAC</b> 4 DRAWER;2 3",1 6",1 9";ADJ DEFIB TRAY;IV POLE (WxDxH) :44x24x52 CM:111.8x61.0x132.1 <b>ELECT:</b> 120V FOR POWER STRIP  -DIMENSIONS LISTED DO NOT INCLUDE SIDE ACCESSORIES, IV POLE HEIGHT, OR DEFIBRILLATOR SHELF HEIGHT	INTERMETRO INDUSTRIES CORP. LECCRP3 LIFELINE	N	OF/OI
6	087848 DEF01	2	<b>DEFIBRILLATOR</b> GUIDELINES 2015 - COMPATIBLE, AED W/MANUAL OVERRIDE (WxDxH) :11x13x8 CM:27.9x33.0x20.3 <b>ELECT:</b> 120VAC;60HZ;2AMPS;SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER	ZOLL MEDICAL CORPORATION 30120000001110012 R SERIES ALS	E	OF/OI
7	092325	2	<b>ASPIRATOR</b> VACUUM;5 SETTINGS;30LPM;BUILT IN AC/DC POWER;LED (WxDxH) :13x7x13 CM:33.0x17.8x33.0 <b>ELECT:</b> 100-240VAC;50/60HZ;RECHARGEABLE 12VDC BATTERY	LAERDAL MEDICAL CORP. 78002001/ACC LSU W/CANISTER	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: STORAGE, EQUIP  
ROOM NUMBER: 6066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	006264	1	<b>NOTE:</b>	ZZZ - NOTE: (U.S.)	I	OF/OI
			- PLACE HOLDER FOR EQUIPMENT			
2	014753 SHV02	1	<b>SHELVING, WIRE</b> MOBILE;CHROME;5 SHELF;4 CASTER STEM BRAKES (WxDxH) :51x25x69 LBS: 60 SPACE: FLOOR CM:129.6x63.5x175.3 KG: 27.30	INTERMETRO INDUSTRIES CORP. 2448NC/63UP/ACC. SUPER ERECTA	N	OF/VI
3	002532 CRT06	1	<b>CART, UTILITY</b> 3 CHROME WIRE SHELVES;5" SWIVEL CASTERS;(2) ONE (WxDxH) :36x24x38 LBS: 39 SPACE: FLOOR CM:91.4x61.0x96.5 KG: 17.70	INTERMETRO INDUSTRIES CORP. 3SPN53DC 3SP SERIES	N	OF/OI
4	058498	1	<b>SPHYGMOMANOMETER, ANEROID, MOBILE</b> 5 LEG STAND;CERTIFIED ACCURATE TO 3MMHG;LATEX (WxDxH) :18x18x36 LBS: 12 SPACE: FLOOR CM:45.7x45.7x91.4 KG: 5.50	WELCH ALLYN, INC. 7670-03	N	OF/OI
5	072678 DOP01	1	<b>DOPPLER, FLOW DETECTOR</b> NON-DIRECTIONAL POCKET DOPPLER W/BUILT-IN (WxDxH) :18x18x36 LBS: 2 SPACE: FLOOR CM:45.7x45.7x91.4 KG: 0.90 <b>ELECT:</b> 9V ALKALINE BATTERY	HUNTLEIGH HEALTHCARE D900-P-USA/EZ8 D900 DOPPLEX	N	OF/OI
6	086490	1	<b>OPHTHALMOSCOPE/OTOSCOPE</b> MOBILE STAND; PANOPTIC OPHTHALMOSCOPE;MACROVIEW (WxDxH) :22x22x54 SPACE: FLOOR CM:55.9x55.9x137.2 <b>ELECT:</b> 100-240V;50/60HZ;.18 AMPS MAX	WELCH ALLYN, INC. 77710-82M/ACC GS 777 SERIES	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: STORAGE, EQUIP  
ROOM NUMBER: 6066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
7	067194	2	<b>I.V. POLE</b> STAINLESS STEEL;5 LEG;3 IN CASTERS;4 HOOK TOP (WxDxH) :23x6x35 LBS: 32 SPACE: FLOOR CM:58.4x15.2x88.9 KG: 14.50	PRYOR PRODUCTS 135	N	OF/OI
8	088643	2	<b>I.V. INFUSION PUMP</b> GENERAL INFUSION PUMP;SINGLE CHANNEL;POLE CLAMP; (WxDxH) :8x8x6 LBS: 10 SPACE: SPACE CM:20.3x20.3x15.2 KG: 4.50 <b>ELECT:</b> 120V;50/60HZ;35W;1A;1 RECHARGEABLE BATTERY LEAD-ACID 6V;POWER CORD:HOSPITAL-GRADE AC CORD 10 FT LONG W/TRANSPARENT PLUG AND RETAINER PLATE	HOSPIRA, INC. 30010 PLUM 360/MEDNET	N	OF/OI
9	087247	2	<b>WALKER, FOLDING, ADULT</b> SIDE WALKER;ALUMINUM;USER WT CAP 300LBS;HT ADJUST (WxDxH) :19x16x36 LBS: 4 SPACE: FLOOR CM:48.3x40.6x91.4 KG: 1.60	PERFORMANCE HEALTHCARE 081561752 HEMI DAYS INVACARE	N	OF/OI
10	066897	2	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0	MEDLINE INDUSTRIES, INC. MDS89668XW	N	OF/OI
11	096810 BRD01	1	<b>BOARD, PATIENT TRANSFER</b> 529 LB CAPACITY;FOR USE IN RADIOLOGY,SURGERY,ICU (WxDxH) :20x70x2 LBS: 8 SPACE: WALL CM:50.8x177.8x5.1 KG: 3.80	SAMARIT 440.0100 HIGHTEC	N	OF/OI
12	078455 BRD03	1	<b>BOARD, PATIENT TRANSFER</b> SURGIBOARD 47X16";TO AND FROM OR TABLE (WxDxH) :16x47x2 SPACE: WALL CM:40.6x119.4x5.1	MCAULEY MEDICAL INC 440.0400 SAMARIT	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL

ROOM NAME: STORAGE, EQUIP

ROOM NUMBER: 6066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
13	055330	1	<b>LIFTER, PATIENT</b> INFLATABLE;NYLON OXFORD;TESTED UPT 1100LBS (WxDxH) :39x70x30 SPACE: FLOOR CM:99.1x177.8x76.2 <b>ELECT:</b> 120V/60HZ/1100 WATTS;GROUNDED;14' POWER CORD	HOVERTECH INTERNATIONAL HJ3902/AIR400G/ACC HOVERJACK	N	OF/OI
14	057580 LFT02	1	<b>CART</b> HOLDS HOVERJACK AIR SUPPLY AND LIFT (WxDxH) :25x18x38 LBS: 26 SPACE: FLOOR CM:63.5x45.7x96.5 KG: 11.80	HOVERTECH INTERNATIONAL HJC-100 HOVER JACK	N	OF/OI
15	073144	1	<b>TRANSILLUMINATOR</b> VEIN LOCATOR;INFRARED LIGHT;HAND HELD;BATTERY (WxDxH) :24x24x36 SPACE: FLOOR CM:61.0x61.0x91.4	BODIFLOW, LLC AV300 ACCUVEIN	N	OF/OI
16	094909 CRT07	1	<b>CART, UTILITY</b> MULTIPURPOSE CART;800LB CAPACITY;5"WHEELS (WxDxH) :21x31x42 LBS: 44 SPACE: FLOOR CM:53.3x78.7x106.7 KG: 20.00	VERSACART EZTOTE 390 101-390	N	OF/OI
17	074251	2	<b>CRUTCHES</b> ALUMINUM;HAND GRIP;I-BEAM ADJUSTMENT;300LB CAP SPACE: SPACE	MEDLINE INDUSTRIES, INC. QUICK FIT MDS80540	N	OF/OI
18	082590	1	<b>WALKER</b> 4-WHEEL WALKER;8" NON-MARRING CASTERS;300LB CAP (WxDxH) :24x26x37 LBS: 20 SPACE: FLOOR CM:61.0x66.0x94.0 KG: 9.10	PERFORMANCE HEALTHCARE 081576974 ROLLATOR WALKER	N	OF/OI





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: STORAGE, EQUIP  
ROOM NUMBER: 6066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
19	091438	1	<b>CHAIR, SHOWER</b> 500LB USER WT CAP;14-19" HT ADJ SEAT;W/KNOCK DOWN (WxDxH) :18x20x30 LBS: 8 SPACE: FLOOR CM:45.7x50.8x76.2 KG: 3.40	DRIVE MEDICAL 12021KD-1 DELUXE BARI	N	OF/OI
20	074071 LFT01	4	<b>LIFTER, PATIENT</b> MOBILE;400LB CAP;ACTIVE STANDING/TRANSFER AID (WxDxH) :29x33x42 LBS: 65 SPACE: FLOOR CM:73.7x83.8x106.7 KG: 29.50 - NEED TO CONFIRM EXISTING	ARJOHUNTLEIGH SARA STEDY NTB2000	N	OF/OI
21	078100	6	<b>LIFTER, PATIENT, ACCESSORY, MOTOR</b> 625LB LIFT MOTOR, 24VDC, CEILING RAIL MOUNTED; (WxDxH) :13x9x6 LBS: 18 SPACE: SPACE CM:33.0x22.9x15.2 KG: 8.20 <b>ELECT:</b> 24VDC, 1.5A	PRISM MEDICAL C625 MOTOR C625 MOTOR	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, NUTRITION  
ROOM NUMBER: 6067

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	095453	1	<b>CART, FOOD</b>	LAKESIDE MFG. CO.	N	OF/OI
	CRT02		SS;6 TRAY CAPACITY;6"LEDGE SPACING;REMOVABLE DOOR (WxDxH) :33x22x46 LBS: 110 SPACE: FLOOR CM:83.8x55.9x116.9 KG: 49.90	SP-6373 ELITE		
			- SOILED TRAY CART			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: ALCOVE, LINEN  
ROOM NUMBER: 6068

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	096814 CRT03	1	<b>CART, LINEN</b> SS;3 FIXED SHELF;6" CASTERS (WxDxH) :62x24x70 LBS: 125 SPACE: FLOOR CM:157.5x61.0x177.8 KG: 56.70	G.S. MANUFACTURING GSM-G160 -	N	OF/OI
2	006698 WST02	1	<b>WASTE RECEPTACLE</b> 44 GALLON;ROUND W/FLAT LID;GRAY CONTAINER;DOLLY (WxDxH) :24x24x37 LBS: 24 SPACE: FLOOR CM:61.0x61.0x94.0 KG: 10.90 SPECIFY COLOR: LID - FOR CLEAN GOWNS	RUBBERMAID COMM. PRODUCTS 2643-60/2645-60/2640 BRUTE	N	OF/OI
3	078147 CAB01	1	<b>CABINET, WARMING</b> MOBILE;2 COMPARTMENTS;20.6CF;GLASS DOORS;SEPARATE (WxDxH) :32x25x74 LBS: 358 SPACE: FLOOR CM:81.3x63.5x188.0 KG: 162.40 <b>ELECT:</b> 120V;50/60HZ;1PH;15A;1.8KW NEMA 5-20P;20A-125V PLUG HOSPITAL GRADE SPECIFY DOOR HINGE;RIGHT OR LEFT - PLACEHOLDER FOR FULL SIZE, SINGLE COMPARTMENT	PEDIGO PRODUCTS, INC. P-2055 P-2055	N	OF/OI
4	002532 CRT06	1	<b>CART, UTILITY</b> 3 CHROME WIRE SHELVES;5" SWIVEL CASTERS;(2) ONE (WxDxH) :36x24x38 LBS: 39 SPACE: FLOOR CM:91.4x61.0x96.5 KG: 17.70	INTERMETRO INDUSTRIES CORP. 3SPN53DC 3SP SERIES	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 6074

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	058112	1	<b>REFRIGERATOR/FREEZER, UPRIGHT</b> 115V;18.2 CF;TOP FREEZER;NON-ICE;LEFT HINGE;WHITE (WxDxH) :30x32x67 LBS: 205 SPACE: FLOOR CM:76.2x81.3x170.2 KG: 93.00 <b>ELECT:</b> 115V;4.5A;60HZ;407 WATTS	SEARS COMMERCIAL SALES KENMORE 46-68972	N	OF/OI
2	087445	1	<b>ICE MAKER W/WATER DISPENSER</b> 12# STORAGE CAP;425#/24HR;LEVER;W/O FILTER (WxDxH) :17x24x33 LBS: 199 SPACE: COUNTER CM:43.2x61.0x83.8 KG: 90.30 <b>ELECT:</b> 115VAC/60HZ/1PH/11A/.8KW 8.5' CORD W/NEMA 5-15 HOSPITAL GRADE PLUG <b>HVAC:</b> 5000 BTU/HR ----- CLEARANCES: VENTILATION 6" TOP & RIGHT SIDES SERVICE 12" TOP <b>PLUMB:</b> REFER TO VENDOR TECHNICAL INFORMATION	FOLLETT CORPORATION 12CI425A-L AIR COOLED	N	OF/CI
3	080483	1	<b>OVEN, MICROWAVE</b> 1000W;S.S. INTERIOR/EXTERIOR (WxDxH) :21x16x13 LBS: 37 SPACE: COUNTER CM:53.3x40.6x33.0 KG: 16.80 <b>ELECT:</b> 120V;14A/11;1.6KW;15AMP NEMA 15-R RECEPTACLE	SHARP ELECTRONICS CORP. R21LCF R21LCF	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 6074

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
4	062450	1	<b>COFFEE BREWER</b> PLUMBED;1 LOWER BREWER/2 UPPER WARMERS;200 OZ CAP. (WxDxH) :9x22x19 LBS: 30 SPACE: COUNTER CM:22.9x55.9x48.3 KG: 13.60 <b>ELECT:</b> 120V/15AMPS;TANK HEATER 1425W;TOTAL WATTS 1800 CORD ATTACHED;REQ.2-WIRES PLUS GROUND SERVICE RATED 120V,SINGLE PHASE,60HZ. <b>PLUMB:</b> 20-90 PSI (138-621 KPA).MACHINES SUPPLIED W/ 1/4" MALE FLARE FITTING - TO INCLUDE HOT WATER TAP	BUNN-O-MATIC CORP. AXIOM-15-3-0000/ACC 38700.0000	L	OF/VI
5	051789	1	<b>TOASTER, COMMERCIAL</b> 2-SLOT;CHROME-PLATED CONSTRUCTION;1 3/8"WIDE SLOTS (WxDxH) :14x9x9 LBS: 5 SPACE: COUNTER CM:35.6x22.9x22.9 KG: 2.30 <b>ELECT:</b> 120 VOLTS/8 AMPS/925 WATTS/PHASE 1	WARING PRODUCTS WCT702 -	N	OF/OI
6	094915	1	<b>DISPENSER, FOOD SERVICE, BEVERAGE</b> SOFT DRINK;4 POST MIXING DISPENSING VALVES;LEG KIT (WxDxH) :11x26x28 LBS: 145 SPACE: COUNTER CM:27.9x66.0x71.1 KG: 65.80 <b>ELECT:</b> 120V;60HZ;7 AMPS <b>PLUMB:</b> SS LABELED INLET SYRUP/WATER;3/8"BARB - COORDINATION REQUIRED FOR SYRUP, CO2, ETC UNDER COUNTER	LANCER CORP. CED 500	L	OF/OI
7	061957	1	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 SPACE: WALL CM:45.7x10.2x22.9	CLINTON INDUSTRIES G-1030 -	N	OF/CI



A division of **Ross&Baruzzini**

712A

**MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT**

-

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 6074

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
8	006677	1	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 LBS: 43 SPACE: FLOOR CM:30.5x53.3x88.9 KG: 19.50 SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI

**END OF ROOM**



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: STORAGE, RT TANK  
ROOM NUMBER: 6080

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	084658 CRT01	3	<b>CART, CYLINDER, MEDICAL GAS</b> 6 CAPACITY; E/D SIZE;W/HEAVY DUTY 5" CASTERS (WxDxH) :11x16x39 LBS: 22 SPACE: FLOOR CM:27.9x40.6x99.1 KG: 10.00	W.T. FARLEY, INC. CR-DC06E-G -	N	OF/OI
2	089776	8	<b>HUMIDIFIER</b> WARMED/HUMIDIFIED RESPIRATORY GASES;2-60 L/MIN (WxDxH) :7x12x7 LBS: 5 SPACE: COUNTER CM:17.8x30.5x17.8 KG: 2.30 <b>ELECT:</b> 100-115V;50/60HZ;2.2 AMPS	FISHER & PAYKEL HEALTHCARE AIRVO 2	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 6721

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	006264	1	<b>NOTE:</b>	ZZZ - NOTE: (U.S.)	I	OF/OI
			- 10 REGULAR ROOMS, 1 ISOLATION, 1 POSITIVE PRESSURE			
2	086193 BED01	12	<b>BED, MED-SURG</b> 500# CAP;IBED UPGRADEABLE;INBED SCALE;ISOFLEX (WxDxH) :43x93x30 SPACE: FLOOR CM:109.2x236.2x76.2	STRYKER CORP/MEDICAL DIV S3 PX2-3005 S3	E	OF/OI
			- *ASSUME RELOCATED FROM RICHARDS (R1, R3, R4, R5). CONFIRMATION NEEDED. CONFIRM O2 TANK HOLDER QUANTITIES			
3	053630	12	<b>BED ACCESSORIES</b> FULLY DYNAMIC;NON-PWR;PRESSURE MANAGEMENT SURFACE (WxDxH) :35x84x7 LBS: 50 SPACE: SPACE CM:88.9x213.4x17.8 KG: 22.70	B.G. INDUSTRIES ACCUMAX QUANTUM A3-117610-384	E	OF/OI
			<b>ELECT:</b> 110 VOLT/PHASE 1 - *ASSUME RELOCATED FROM RICHARDS (R1, R3, R4, R5). CONFIRMATION NEEDED.			
4	096840 LFT05	12	<b>LIFTER, PATIENT, ACCESSORY, RAIL</b> 10'STRAIGHT;FOR USE WITH C1000 MOTOR SPACE: CEILING	PRISM MEDICAL 10' STRAIGHT	N	OF/VI





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 6721

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
5	086984 MON06	12	<b>MONITOR, VITAL SIGNS</b> 8.4"LCD SCREEN;NBP;SPO2;TYMPANIC TEMP;WIRELESS (WxDxH) :11x6x9 LBS: 7 SPACE: WALL CM:27.9x15.2x22.9 KG: 3.10 <b>ELECT:</b> 120 VOLT/60 HZ BATTERY CHARGER LITHIUM ION BATTERY 10.8-11.1V DATA OUTPUT:HL7 FORMAT, VIA ETHERNET PORT SERIAL DATA	PHILIPS MEDICAL SYSTEMS 863283 SURESIGNS VS4	N	OF/OI
6	093331 WCH01	12	<b>WALL CHANNEL</b> WALL CHANNEL;SEISMIC;19" SPACE: WALL	PHILIPS MEDICAL SYSTEMS MXU175	N	OF/CI
7	077835 MNT01	12	<b>MONITOR ACCESSORY, MOUNTING ASSEMBLY</b> VARIABLE HEIGHT MOUNT;TILT/SWIVEL ADJUSTMENT; SPACE: WALL	GCX CORPORATION WA-0011-27 VHM-25	N	OF/OI
8	028803	12	<b>STAND, MAYO</b> MOBILE;CHROME;HAND HEIGHT ADJUSTABLE 30-40 IN (WxDxH) :13x20x40 LBS: 22 SPACE: FLOOR CM:33.0x50.8x101.6 KG: 10.00 - PLACEHOLDER. SPECS TO BE DETERMINED. MAY NOT NEED 1 PER ROOM.	PEDIGO PRODUCTS, INC. P-65 MOBILE	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 6721

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
9	067194	12	<b>I.V. POLE</b> STAINLESS STEEL;5 LEG;3 IN CASTERS;4 HOOK TOP (WxDxH) :23x6x35 LBS: 32 SPACE: FLOOR CM:58.4x15.2x88.9 KG: 14.50 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	PRYOR PRODUCTS 135	N	OF/OI
10	088643	12	<b>I.V. INFUSION PUMP</b> GENERAL INFUSION PUMP;SINGLE CHANNEL;POLE CLAMP; (WxDxH) :8x8x6 LBS: 10 SPACE: SPACE CM:20.3x20.3x15.2 KG: 4.50 <b>ELECT:</b> 120V;50/60HZ;35W;1A;1 RECHARGEABLE BATTERY LEAD-ACID 6V;POWER CORD:HOSPITAL-GRADE AC CORD 10 FT LONG W/TRANSPARENT PLUG AND RETAINER PLATE - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	HOSPIRA, INC. 30010 PLUM 360/MEDNET	N	OF/OI
11	096809	12	<b>PUMP, COMPRESSION</b> DVT;LCD SCREEN;ONE BUTTON START;CARRY HANDLE (WxDxH) :9x8x10 LBS: 9 SPACE: SPACE CM:22.9x20.3x25.4 KG: 4.10 <b>ELECT:</b> 120V/60HZ - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	ARJOHUNTLEIGH ACS900 FLOWTRON	L	OF/OI
12	066897	12	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	MEDLINE INDUSTRIES, INC. MDS89668XW	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 6721

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
13	087247	12	<b>WALKER, FOLDING, ADULT</b> SIDE WALKER;ALUMINUM;USER WT CAP 300LBS;HT ADJUST (WxDxH) :19x16x36 LBS: 4 SPACE: FLOOR CM:48.3x40.6x91.4 KG: 1.60 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	PERFORMANCE HEALTHCARE 081561752 HEMI DAYS INVACARE	N	OF/OI
14	063659	12	<b>TABLE, OVERBED</b> SINGLE TOP;W/VANITY;LOW PROFILE U-SHAPED BASE (WxDxH) :21x32x45 LBS: 53 SPACE: FLOOR CM:53.3x81.3x114.3 KG: 24.10 SPECIFY FINISH SPECIFY FINISH: BASE SPECIFY FINISH: EDGE	STRYKER CORP/MEDICAL DIV 3150-000-200 TRU-FIT	E	OF/OI
15	080275	24	<b>REGULATOR, SUCTION, CONTINUOUS &amp; INTERMITTENT</b> 0-200MMHG W/SELECTOR VALVE (OFF,INT,CONT);CANISTER (WxDxH) :4x6x4 LBS: 1 SPACE: WALL CM:10.2x15.2x10.2 KG: 0.50	BOEHRINGER LABORATORIES, INC. 3804/1496/2469 PLATINUM SERIES	E	OF/OI
16	055048	12	<b>FLOWMETER, OXYGEN</b> W/DISS ADAPTOR  SPACE: SPACE  CONFIRM MEDICAL GAS CONNECTION	AMVEX, DIVISION OF OHIO MEDICAL FM15ODH -	E	OF/OI
17	062114 DSP01	12	<b>DISPENSER</b> EMEBAG DISPENSER (WxDxH) :7x7x7 SPACE: WALL CM:17.8x17.8x17.8	CENTURION MEDICAL PRODUCTS EMED200 EMEBAG	N	OF/CI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 6721

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
18	091196 DSP03	12	<b>DISPENSER, MASK</b> PLEXIGLAS DISPENSER FOR SURGICAL FACE MASK BOXES (WxDxH) :9x6x9 LBS: 2 SPACE: WALL CM:22.9x15.2x22.9 KG: 0.90	MEDLINE INDUSTRIES, INC. NONFM122 NONFM122	N	OF/OI
19	087029 DSL01	12	<b>DISPOSAL CONTAINER, SHARPS W/BRACKET</b> LOCKING WALL CABINET FOR SHARPS CONTAINERS;2 GAL (WxDxH) :15x6x14 SPACE: WALL CM:38.1x15.2x35.6	STERICYCLE OC-02-2004 BIO SYSTEMS	L	OF/CI
20	061957 GLV02	12	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 SPACE: WALL CM:45.7x10.2x22.9	CLINTON INDUSTRIES G-1030 -	N	OF/CI
21	058768 HMP01	12	<b>HAMPER, LINEN</b> W/LID;FOOT PEDAL;STEEL FRAME;POWDER COAT FINISH (WxDxH) :21x22x33 SPACE: FLOOR CM:53.3x55.9x83.8 - INSTALLED IN CASEWORK	CENTURION MEDICAL PRODUCTS CX302 MEDIUM	N	OF/OI
22	078043 HMP02	12	<b>HAMPER, LINEN</b> W/LID;FOOT PEDAL OPERATED;TUBULAR CHROME FRAME; (WxDxH) :19x20x26 SPACE: FLOOR CM:48.3x50.8x66.0 - FOR YELLOW BAG CHEMO. INSTALLED IN CASEWORK	CENTURION MEDICAL PRODUCTS CENTURION CX502	N	OF/OI
23	047636	12	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60	RUBBERMAID COMM. PRODUCTS 2544 40 QT	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 6721

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
24	085429 RLS01	24	<b>RAIL SYSTEM</b> 7"2" VERTICAL RAIL SYSTEM;SURFACE MOUNT;1 TIER SPACE: WALL	MODULAR SERVICES COMPANY 530-4522	N	CF/CI
			- WITH BASKET, SHELF, AND OTHER ACCESSORIES TO BE DETERMINED.			
25	504469	12	<b>CLOCK</b>	ZZZ - GENERIC DESCRIPTIONS	N	OF/OI
			- SPECIFIED BY OTHERS			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E6-2 ONCOLOGY-SURGICAL  
ROOM NAME: TOILET, PATIENT (QTY 12)  
ROOM NUMBER: 6721A

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	047636	12	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60	RUBBERMAID COMM. PRODUCTS 2544 40 QT	N	OF/OI

END OF ROOM  
END OF DEPARTMENT



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: CORRIDOR  
ROOM NUMBER: 7004

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	093412 MVD01	4	<b>MONITOR, VIDEO</b> MXU0299: DISPLAY FLAT PANEL NON-TOUCH PIIC SPACE: SPACE	PHILIPS MEDICAL SYSTEMS MXU0299 866390	N	OF/OI
2	093331 WCH01	4	<b>WALL CHANNEL</b> WALL CHANNEL;SEISMIC;19" SPACE: WALL	PHILIPS MEDICAL SYSTEMS MXU175	N	OF/CI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: PATIENT CARE STATION (QTY 5)  
ROOM NUMBER: 7005

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	087769 MON01	4	<b>MONITOR, CENTRAL STATION</b> PIIC IX SURVEILLANCE CENTRAL MONITORING;19" TOUCH (WxDxH) :48x16x16 SPACE: COUNTER CM:121.9x40.6x40.6 <b>ELECT:</b> CPU: 115 VAC;60 HZ;5.2 AMPS LCD DISPLAY: 115 VAC;60 HZ;48W TYP. UPS: 115 VAC;60 HZ;5.6 AMPS PRINTER: 117 VAC;60 HZ;5.3 AMPS <b>HVAC:</b> MUST MAINTAIN TEMP AT 62-82 DEGREES F AND NON-CONDENSING RELATIVE HUMIDITY AT 30-60% CPU: 215 BTU/HR UPS: 437 BTU/HR (MAX) DATA CONNECTION REQUIRED, SEE TECH SHEET	PHILIPS MEDICAL SYSTEMS INTELLIVUE INFO CENTER IX	N	OF/OI
2	505366	1	<b>MONITORING SYSTEM</b>	ZZZ - MONITORING	N	OF/OI
			- ALLOWANCE FOR MONITORING INFRASTRUCTURE			
3	086723 ANA01	3	<b>LAB ANALYZER, GLUCOSE</b> BLOOD GLUCOSE & KETONE MONITORING;BARCODE READER (WxDxH) :3x2x8 LBS: 1 SPACE: SPACE CM:7.6x5.1x20.3 KG: 0.50 <b>ELECT:</b> 2 AA BATTERIES (ALKALINE, LITHIUM, OR NICAD)	ABBOTT DIAGNOSTICS 71411-70 PRECISION PRO	N	OF/OI

END OF ROOM





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, EQ (QTY 2)  
ROOM NUMBER: 7011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	006264	1	<b>NOTE:</b>	ZZZ - NOTE: (U.S.)	I	OF/OI
			- PLACE HOLDER FOR EQUIPMENT			
2	079188 SHV01	2	<b>SHELVING, WIRE</b> 5 SHELVES;STATIONARY;CHROME;SHELF INLAY (WxDxH) :24x24x74 SPACE: FLOOR CM:61.0x61.0x188.0	INTERMETRO INDUSTRIES CORP. A2424NC/74P A2424NC	I	OF/VI
3	002532 CRT06	2	<b>CART, UTILITY</b> 3 CHROME WIRE SHELVES;5" SWIVEL CASTERS;(2) ONE (WxDxH) :36x24x38 LBS: 39 SPACE: FLOOR CM:91.4x61.0x96.5 KG: 17.70	INTERMETRO INDUSTRIES CORP. 3SPN53DC 3SP SERIES	I	OF/OI
4	058498	2	<b>SPHYGMOMANOMETER, ANEROID, MOBILE</b> 5 LEG STAND;CERTIFIED ACCURATE TO 3MMHG;LATEX (WxDxH) :18x18x36 LBS: 12 SPACE: FLOOR CM:45.7x45.7x91.4 KG: 5.50	WELCH ALLYN, INC. 7670-03	I	OF/OI
5	072678 DOP01	2	<b>DOPPLER, FLOW DETECTOR</b> NON-DIRECTIONAL POCKET DOPPLER W/BUILT-IN (WxDxH) :18x18x36 LBS: 2 SPACE: FLOOR CM:45.7x45.7x91.4 KG: 0.90 <b>ELECT:</b> 9V ALKALINE BATTERY	HUNTLEIGH HEALTHCARE D900-P-USA/EZ8 D900 DOPPLEX	I	OF/OI
6	086490	1	<b>OPHTHALMOSCOPE/OTOSCOPE</b> MOBILE STAND; PANOPTIC OPHTHALMOSCOPE;MACROVIEW (WxDxH) :22x22x54 SPACE: FLOOR CM:55.9x55.9x137.2 <b>ELECT:</b> 100-240V;50/60HZ;.18 AMPS MAX	WELCH ALLYN, INC. 77710-82M/ACC GS 777 SERIES	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, EQ (QTY 2)  
ROOM NUMBER: 7011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
7	067194	4	<b>I.V. POLE</b> STAINLESS STEEL;5 LEG;3 IN CASTERS;4 HOOK TOP (WxDxH) :23x6x35 LBS: 32 SPACE: FLOOR CM:58.4x15.2x88.9 KG: 14.50	PRYOR PRODUCTS 135	I	OF/OI
8	088643	4	<b>I.V. INFUSION PUMP</b> GENERAL INFUSION PUMP;SINGLE CHANNEL;POLE CLAMP; (WxDxH) :8x8x6 LBS: 10 SPACE: SPACE CM:20.3x20.3x15.2 KG: 4.50 <b>ELECT:</b> 120V;50/60HZ;35W;1A;1 RECHARGEABLE BATTERY LEAD-ACID 6V;POWER CORD:HOSPITAL-GRADE AC CORD 10 FT LONG W/TRANSPARENT PLUG AND RETAINER PLATE	HOSPIRA, INC. 30010 PLUM 360/MEDNET	I	OF/OI
9	087247	4	<b>WALKER, FOLDING, ADULT</b> SIDE WALKER;ALUMINUM;USER WT CAP 300LBS;HT ADJUST (WxDxH) :19x16x36 LBS: 4 SPACE: FLOOR CM:48.3x40.6x91.4 KG: 1.60	PERFORMANCE HEALTHCARE 081561752 HEMI DAYS INVACARE	I	OF/OI
10	066897	4	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0	MEDLINE INDUSTRIES, INC. MDS89668XW	I	OF/OI
11	096810 BRD01	2	<b>BOARD, PATIENT TRANSFER</b> 529 LB CAPACITY;FOR USE IN RADIOLOGY,SURGERY,ICU (WxDxH) :20x70x2 LBS: 8 SPACE: WALL CM:50.8x177.8x5.1 KG: 3.80	SAMARIT 440.0100 HIGHTEC	I	OF/OI
12	078455 BRD03	2	<b>BOARD, PATIENT TRANSFER</b> SURGIBOARD 47X16";TO AND FROM OR TABLE (WxDxH) :16x47x2 SPACE: WALL CM:40.6x119.4x5.1	MCAULEY MEDICAL INC 440.0400 SAMARIT	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, EQ (QTY 2)  
ROOM NUMBER: 7011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
13	055330	1	<b>LIFTER, PATIENT</b> INFLATABLE;NYLON OXFORD;TESTED UPT 1100LBS (WxDxH) :39x70x30 SPACE: FLOOR CM:99.1x177.8x76.2 <b>ELECT:</b> 120V/60HZ/1100 WATTS;GROUNDED;14' POWER CORD	HOVERTECH INTERNATIONAL HJ3902/AIR400G/ACC HOVERJACK	I	OF/OI
14	057580 LFT02	1	<b>CART</b> HOLDS HOVERJACK AIR SUPPLY AND LIFT (WxDxH) :25x18x38 LBS: 26 SPACE: FLOOR CM:63.5x45.7x96.5 KG: 11.80	HOVERTECH INTERNATIONAL HJC-100 HOVER JACK	I	OF/OI
15	050708 MON03	1	<b>MONITOR, PHYSIOLOGICAL, TRANSPORT</b> BIPHASIC DEFIBRILLATOR;EXTERNAL PACING;PRINTER (WxDxH) :11x9x11 LBS: 18 SPACE: SPACE CM:27.9x22.9x27.9 KG: 8.20 <b>ELECT:</b> 120V;60HZ;	ZOLL MEDICAL CORPORATION 43510711100124010/AC M CCT/BIPHASIC	I	OF/OI
16	079214 MON04	1	<b>MONITOR, PHYSIOLOGICAL, TRANSPORT</b> TRANSPORT MONITOR WITH MMS, TO INCLUDE: (WxDxH) :13x8x12 LBS: 13 SPACE: SPACE CM:33.0x20.3x30.5 KG: 5.90 <b>ELECT:</b> 100-240V, 50/60HZ, 70W AVERAGE, 1.2AMP <b>PLUMB:</b> NO PLUMBING REQUIRED	PHILIPS MEDICAL SYSTEMS 866060 MX400	I	OF/OI
17	073144	1	<b>TRANSILLUMINATOR</b> VEIN LOCATOR;INFRARED LIGHT;HAND HELD;BATTERY (WxDxH) :24x24x36 SPACE: FLOOR CM:61.0x61.0x91.4	BODIFLOW, LLC AV300 ACCUVEIN	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, EQ (QTY 2)  
ROOM NUMBER: 7011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
18	094909 CRT07	2	<b>CART, UTILITY</b> MULTIPURPOSE CART;800LB CAPACITY;5"WHEELS (WxDxH) :21x31x42 LBS: 44 SPACE: FLOOR CM:53.3x78.7x106.7 KG: 20.00	VERSACART EZTOTE 390 101-390	E	OF/OI
19	074251	4	<b>CRUTCHES</b> ALUMINUM;HAND GRIP;I-BEAM ADJUSTMENT;300LB CAP SPACE: SPACE	MEDLINE INDUSTRIES, INC. QUICK FIT MDS80540	I	OF/OI
20	070331 SCL01	4	<b>SCALE, STAND-ON</b> PORTABLE;DIGITAL READOUT;HANDRAIL/HEIGHT GAUGE (WxDxH) :26x26x78 LBS: 62 SPACE: FLOOR CM:66.0x66.0x198.1 KG: 28.10 <b>ELECT:</b> 120V; 6-D BATTERIES	SCALE-TRONIX DIV. OF WELCH ALLYN 5002/845010/845233	E	OF/OI
21	082590	2	<b>WALKER</b> 4-WHEEL WALKER;8" NON-MARRING CASTERS;300LB CAP (WxDxH) :24x26x37 LBS: 20 SPACE: FLOOR CM:61.0x66.0x94.0 KG: 9.10	PERFORMANCE HEALTHCARE 081576974 ROLLATOR WALKER	I	OF/OI
22	091438	1	<b>CHAIR, SHOWER</b> 500LB USER WT CAP;14-19" HT ADJ SEAT;W/KNOCK DOWN (WxDxH) :18x20x30 LBS: 8 SPACE: FLOOR CM:45.7x50.8x76.2 KG: 3.40	DRIVE MEDICAL 12021KD-1 DELUXE BARI	E	OF/OI
23	074071 LFT01	1	<b>LIFTER, PATIENT</b> MOBILE;400LB CAP;ACTIVE STANDING/TRANSFER AID (WxDxH) :29x33x42 LBS: 65 SPACE: FLOOR CM:73.7x83.8x106.7 KG: 29.50	ARJOHUNTLEIGH SARA STEDY NTB2000	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, EQ (QTY 2)  
ROOM NUMBER: 7011

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
24	089538	2	<b>EVACUATION EQUIPMENT, PATIENT</b> 36" STANDARD EVACUATION SLED  SPACE: SPACE	MED SLED MS36 STANDARD	E	OF/OI
25	089251 CRT05	1	<b>CART, RESUSCITATION, CARDIAC</b> 4 DRAWER;2 3",1 6",1 9";ADJ DEFIB TRAY;IV POLE (WxDxH) :44x24x52 CM:111.8x61.0x132.1 <b>ELECT:</b> 120V FOR POWER STRIP  -DIMENSIONS LISTED DO NOT INCLUDE SIDE ACCESSORIES, IV POLE HEIGHT, OR DEFIBRILLATOR SHELF HEIGHT  SPACE: FLOOR	INTERMETRO INDUSTRIES CORP. LECCRP3 LIFELINE	N	OF/OI
26	087848 DEF01	1	<b>DEFIBRILLATOR</b> GUIDELINES 2015 - COMPATIBLE, AED W/MANUAL OVERRIDE (WxDxH) :11x13x8 LBS: 14 SPACE: COUNTER CM:27.9x33.0x20.3 KG: 6.40 <b>ELECT:</b> 120VAC;60HZ;2AMPS;SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER	ZOLL MEDICAL CORPORATION 30120000001110012 R SERIES ALS	N	OF/OI
27	092325	1	<b>ASPIRATOR</b> VACUUM;5 SETTINGS;30LPM;BUILT IN AC/DC POWER;LED (WxDxH) :13x7x13 LBS: 9 SPACE: SPACE CM:33.0x17.8x33.0 KG: 4.10 <b>ELECT:</b> 100-240VAC;50/60HZ;RECHARGEABLE 12VDC BATTERY	LAERDAL MEDICAL CORP. 78002001/ACC LSU W/CANISTER	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, LINEN (QTY 2)  
ROOM NUMBER: 7012

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	096814 CRT03	2	<b>CART, LINEN</b> SS;3 FIXED SHELF;6" CASTERS (WxDxH) :62x24x70 LBS: 125 SPACE: FLOOR CM:157.5x61.0x177.8 KG: 56.70	G.S. MANUFACTURING GSM-G160 -	N	OF/OI
2	006698 WST02	1	<b>WASTE RECEPTACLE</b> 44 GALLON;ROUND W/FLAT LID;GRAY CONTAINER;DOLLY (WxDxH) :24x24x37 LBS: 24 SPACE: FLOOR CM:61.0x61.0x94.0 KG: 10.90 SPECIFY COLOR: LID - FOR CLEAN GOWNS	RUBBERMAID COMM. PRODUCTS 2643-60/2645-60/2640 BRUTE	N	OF/OI
3	078147 CAB01	2	<b>CABINET, WARMING</b> MOBILE;2 COMPARTMENTS;20.6CF;GLASS DOORS;SEPARATE (WxDxH) :32x25x74 LBS: 358 SPACE: FLOOR CM:81.3x63.5x188.0 KG: 162.40 <b>ELECT:</b> 120V;50/60HZ;1PH;15A;1.8KW NEMA 5-20P;20A-125V PLUG HOSPITAL GRADE SPECIFY DOOR HINGE;RIGHT OR LEFT - PLACEHOLDER FOR FULL SIZE, SINGLE COMPARTMENT	PEDIGO PRODUCTS, INC. P-2055 P-2055	N	OF/OI
4	002532 CRT06	2	<b>CART, UTILITY</b> 3 CHROME WIRE SHELVES;5" SWIVEL CASTERS;(2) ONE (WxDxH) :36x24x38 LBS: 39 SPACE: FLOOR CM:91.4x61.0x96.5 KG: 17.70	INTERMETRO INDUSTRIES CORP. 3SPN53DC 3SP SERIES	N	OF/OI

END OF ROOM



A division of **Ross&Baruzzini**

712A

**MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT**

-

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, NUTRITION (QTY 2)  
ROOM NUMBER: 7013

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	095453	2	<b>CART, FOOD</b>	LAKESIDE MFG. CO.	E	OF/OI
	CRT02		SS;6 TRAY CAPACITY;6"LEDGE SPACING;REMOVABLE DOOR (WxDxH) :33x22x46 LBS: 110 SPACE: FLOOR CM:83.8x55.9x116.9 KG: 49.90	SP-6373 ELITE		
			- SOILED TRAY CART			

**END OF ROOM**



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: SOILED HOLDING  
ROOM NUMBER: 7015

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	094910 COM01	1	<b>COMPACTOR, WASTE</b> 20:1 COMPACT RATION;COMMERCIAL;ELECTROHYDRAULIC (WxDxH) :43x31x90 LBS: 1210 SPACE: FLOOR CM:109.2x78.7x228.6 KG: 548.80 <b>ELECT:</b> 115V;60HZ;15 AMP;SINGLE PHASE;DEDICATED CIRCUIT - ECO-PACK	FOUNTAIN INDUSTRIES ECO-PACK 3600	N	OF/OI
2	079188 SHV01	1	<b>SHELVING, WIRE</b> 5 SHELVES;STATIONARY;CHROME;SHELF INLAY (WxDxH) :24x24x74 SPACE: FLOOR CM:61.0x61.0x188.0	INTERMETRO INDUSTRIES CORP. A2424NC/74P A2424NC	N	OF/OI
3	073358 TRK01	1	<b>TRUCK/BIN, LINEN</b> HORIZONTAL SHELVES;POLYETHYLENE;IMPACT RESISTANT (WxDxH) :48x29x64 SPACE: FLOOR CM:121.9x73.7x162.6	MEESE ORBITRON DUNNE CO. POLY TRUX 90P	N	OF/OI
4	066897	1	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0	MEDLINE INDUSTRIES, INC. MDS89668XW	N	OF/OI
5	096813	1	<b>WASTE RECEPTACLE, STEP-ON</b> 12 GAL;WHITE;FOOT PEDAL OPERATED;FIRE SAFE (WxDxH) :12x12x23 SPACE: FLOOR CM:30.5x30.5x58.4	RUBBERMAID COMM. PRODUCTS RCPST12ERBWHI DEFENDERS	N	OF/OI





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: SOILED HOLDING  
ROOM NUMBER: 7015

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
6	071200	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> 9 GALLON;CLEAR SLIDE TOP;RECYCLED PLASTIC-BLACK (WxDxH) :18x12x19 SPACE: SPACE CM:45.7x30.5x48.3 SPECIFY ACCESSORIES AS SEPARATE LINE ITEMS	BECTON DICKINSON RCRA HAZARDOUS WASTE 305069	N	OF/OI
7	070664	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> CHEMOTHERAPY;PLASTIC;18 GAL;YELLOW;SLIDING LID (WxDxH) :18x13x26 LBS: 8 SPACE: FLOOR CM:45.7x33.0x66.0 KG: 3.60	KENDALL HEALTHCARE, COVIDIEN 8939 CHEMO DISPOSAL	N	OF/OI
8	089948	1	<b>WASTE RECEPTACLE, RECYCLING</b> 23 GALLON;DARK BLUE;RECYCLE SYMBOL;W/ LID (WxDxH) :20x11x35 SPACE: FLOOR CM:50.8x27.9x88.9	RUBBERMAID COMM. PRODUCTS H-1385BLU SLIMJIM 3540-07	N	OF/OI
9	079178	1	<b>WASTE RECEPTACLE</b> CONFIDENTIAL;SIDE PAPER SLOT;KEYPAD LOCK;SHIPS (WxDxH) :11x24x25 LBS: 30 SPACE: FLOOR CM:27.9x61.0x63.5 KG: 13.60	RUBBERMAID COMM. PRODUCTS 9W25 SLIM JIM	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: MEDICATION  
ROOM NUMBER: 7016

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	072239	1	<b>DISPENSER, MEDICATION, AUTOMATED</b> 14.75"LCD MONITOR;KEYBOARD;TOUCHPAD;BIO-ID SCANNER (WxDxH) :23x27x55 LBS: 166 SPACE: FLOOR CM:58.4x68.6x139.7 KG: 75.30 <b>ELECT:</b> 120V;60HZ;3 AMPS MAX EMERGENCY POWER REQUIRED <b>HVAC:</b> 409 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET - NEED TO UPDATE TO ES SERIES	BD CAREFUSION PYXIS MEDSTATION ES 6 DRAWER MAIN	N	OF/VI
2	067803	1	<b>DISPENSER, MEDICATION, AUTOMATED</b> DOUBLE COLUMN AUXILIARY;8 DOORS (WxDxH) :52x28x80 LBS: 314 SPACE: FLOOR CM:132.1x71.1x203.2 KG: 142.70 <b>ELECT:</b> 120V;60HZ;140 WATTS;DATA DROP REQUIRED <b>HVAC:</b> 362 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET	BD CAREFUSION PYXIS DOUBLE COLUMN AUX MEDSTATION 4000	N	OF/VI
3	072243	1	<b>DISPENSER, MEDICATION, AUTOMATED</b> SINGLE COLUMN;1 STORAGE SHELF PER DOOR;INTERIOR (WxDxH) :31x28x80 LBS: 314 SPACE: FLOOR CM:78.7x71.1x203.2 KG: 142.40 <b>ELECT:</b> 120V;60HZ;3 AMPS;EMERGENCY POWER REQUIRED <b>HVAC:</b> 222 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET	BD CAREFUSION PYXIS FOUR DOOR AUXILIARY MEDSTATION ES	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: MEDICATION  
ROOM NUMBER: 7016

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
4	096490 REF09	2	<b>REFRIGERATOR, UNDERCOUNTER, MEDICAL GRADE</b> 4.5 CU.FT CAPACITY;PROGRAMMED AUTOMATIC DEFROST (WxDxH) :24x27x34 LBS: 175 SPACE: FLOOR CM:61.0x68.6x86.4 KG: 79.50 <b>ELECT:</b> 115V;60HZ; <b>HVAC:</b> BTU 1596 W/H - STACKED UNITS;1 COMPARTMENT HAZARDOUS, 1 NON-HAZARDOUS	FOLLETT CORPORATION REF5P-00-00/ACC PERFORMANCE +	N	OF/OI
5	096211	1	<b>BRACKET</b> FOR STACKING PERFORMANCE PLUS W/PERFORMANCE PLUS LBS: 8 SPACE: SPACE KG: 3.60	FOLLETT CORPORATION STACKING KIT 01054006	N	OF/OI
6	053907 MED01	2	<b>DISPENSER, MEDICATION, ACCESSORY</b> MED DISPENSING UNIT REFRIGERATOR CONTROL (WxDxH) :3x6x7 LBS: 4 SPACE: SPACE CM:7.6x15.2x17.8 KG: 1.80 <b>ELECT:</b> POWERED FROM MAIN DISPENSER	BD CAREFUSION PYXIS REMOTE MANAGER MSRM	N	OF/VI
7	506535 OSS01	1	<b>OPEN SUPPLY SYSTEM</b>  - PAREX SUPPLY SYSTEM, WEIGHT BASED -PROVIDE BACKING ON ALL WALLS WITH NO OUTLETS EXTENDING BEYOND 12" AFF. REQUIRES POWER AND DATA AT CONTROL LOCATION.	ZZZ - GENERIC DESCRIPTIONS	N	OF/VI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: MEDICATION  
ROOM NUMBER: 7016

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
8	077862 BIN01	10	<b>BIN, ACCESSORY, PANEL, LOUVERED</b> LOUVERED PANEL;36";STONE (WxDxH) :36x1x19 CM:91.4x2.5x48.3 SPACE: WALL	PAR EXCELLENCE SYSTEMS, INC. 30-636	N	OF/CI
9	077862 BIN01	2	<b>BIN, ACCESSORY, PANEL, LOUVERED</b> LOUVERED PANEL;36";STONE (WxDxH) :36x1x19 CM:91.4x2.5x48.3 - FOR HAZARDOUS CHEMO BIN SPACE: WALL	PAR EXCELLENCE SYSTEMS, INC. 30-636	N	OF/CI
10	093412 MVD01	1	<b>MONITOR, VIDEO</b> MXU0299: DISPLAY FLAT PANEL NON-TOUCH PIIC SPACE: SPACE	PHILIPS MEDICAL SYSTEMS MXU0299 866390	N	OF/OI
			- TO INCLUDE SPEAKERS			
11	055704 EWS01	1	<b>ENCLOSURE, WALL, SHARPS</b> 2 OR 5 QT WALL ENCLOSURE (WxDxH) :13x6x12 CM:33.0x15.2x30.5 SPACE: WALL	KENDALL HEALTHCARE, COVIDIEN 85161H IN-ROOM	L	OF/CI
12	093535 DSL02	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> 2GAL;BLACK;WHITE HINGED LID W/PORT;LEAK RESISTANT (WxDxH) :12x10x12 CM:30.5x25.4x30.5 SPACE: WALL	KENDALL HEALTHCARE, COVIDIEN 8602RC/8963 8602RC/8963	N	OF/OI
13	061957 GLV02	2	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 CM:45.7x10.2x22.9 SPACE: WALL	CLINTON INDUSTRIES G-1030 -	N	OF/CI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: MEDICATION  
ROOM NUMBER: 7016

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
14	006677	1	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 LBS: 43 SPACE: FLOOR CM:30.5x53.3x88.9 KG: 19.50 SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI
15	047636	2	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60 - PLACEHOLDER FOR RED AND BLACK BINS. NEED VENDOR AND MODEL.	RUBBERMAID COMM. PRODUCTS 2544 40 QT	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 7018

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	058112 REF02	1	<b>REFRIGERATOR/FREEZER, UPRIGHT</b> 115V;18.2 CF;TOP FREEZER;NON-ICE;LEFT HINGE;WHITE (WxDxH) :30x32x67 LBS: 205 SPACE: FLOOR CM:76.2x81.3x170.2 KG: 93.00 <b>ELECT:</b> 115V;4.5A;60HZ;407 WATTS	SEARS COMMERCIAL SALES KENMORE 46-68972	N	OF/OI
2	087445 ICE01	1	<b>ICE MAKER W/WATER DISPENSER</b> 12# STORAGE CAP;425#/24HR;LEVER;W/O FILTER (WxDxH) :17x24x33 LBS: 199 SPACE: COUNTER CM:43.2x61.0x83.8 KG: 90.30 <b>ELECT:</b> 115VAC/60HZ/1PH/11A/.8KW 8.5' CORD W/NEMA 5-15 HOSPITAL GRADE PLUG <b>HVAC:</b> 5000 BTU/HR  CLEARANCES: VENTILATION 6" TOP & RIGHT SIDES SERVICE 12" TOP <b>PLUMB:</b> REFER TO VENDOR TECHNICAL INFORMATION	FOLLETT CORPORATION 12CI425A-L AIR COOLED	N	OF/CI
3	080483 OVE01	1	<b>OVEN, MICROWAVE</b> 1000W;S.S. INTERIOR/EXTERIOR (WxDxH) :21x16x13 LBS: 37 SPACE: COUNTER CM:53.3x40.6x33.0 KG: 16.80 <b>ELECT:</b> 120V;14A/11;1.6KW;15AMP NEMA 15-R RECEPTACLE	SHARP ELECTRONICS CORP. R21LCF R21LCF	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 7018

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
4	062450	1	<b>COFFEE BREWER</b> PLUMBED;1 LOWER BREWER/2 UPPER WARMERS;200 OZ CAP. (WxDxH) :9x22x19 LBS: 30 SPACE: COUNTER CM:22.9x55.9x48.3 KG: 13.60 <b>ELECT:</b> 120V/15AMPS;TANK HEATER 1425W;TOTAL WATTS 1800 CORD ATTACHED;REQ.2-WIRES PLUS GROUND SERVICE RATED 120V,SINGLE PHASE,60HZ. <b>PLUMB:</b> 20-90 PSI (138-621 KPA).MACHINES SUPPLIED W/ 1/4" MALE FLARE FITTING - TO INCLUDE HOT WATER TAP	BUNN-O-MATIC CORP. AXIOM-15-3-0000/ACC 38700.0000	L	OF/VI
5	051789	1	<b>TOASTER, COMMERCIAL</b> 2-SLOT;CHROME-PLATED CONSTRUCTION;1 3/8"WIDE SLOTS (WxDxH) :14x9x9 LBS: 5 SPACE: COUNTER CM:35.6x22.9x22.9 KG: 2.30 <b>ELECT:</b> 120 VOLTS/8 AMPS/925 WATTS/PHASE 1	WARING PRODUCTS WCT702 -	N	OF/OI
6	094915	1	<b>DISPENSER, FOOD SERVICE, BEVERAGE</b> SOFT DRINK;4 POST MIXING DISPENSING VALVES;LEG KIT (WxDxH) :11x26x28 LBS: 145 SPACE: COUNTER CM:27.9x66.0x71.1 KG: 65.80 <b>ELECT:</b> 120V;60HZ;7 AMPS <b>PLUMB:</b> SS LABELED INLET SYRUP/WATER;3/8"BARB - COORDINATION REQUIRED FOR SYRUP, CO2, ETC UNDER COUNTER	LANCER CORP. CED 500	L	OF/OI
7	061957	1	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 SPACE: WALL CM:45.7x10.2x22.9	CLINTON INDUSTRIES G-1030 -	N	OF/CI



A division of **Ross&Baruzzini**

712A

**MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT**

-

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 7018

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
8	006677	1	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 LBS: 43 SPACE: FLOOR CM:30.5x53.3x88.9 KG: 19.50 SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI

**END OF ROOM**





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, CART  
ROOM NUMBER: 7022

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - IV CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	I	OF/OI
2	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - RT CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	I	OF/OI
3	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - FOR LINEN	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	I	OF/OI
4	092750 MON05	1	<b>MONITOR, VITAL SIGNS</b> NIBP;SPO2;TEMPERATURE;RECORDER;ADULT/PEDIATRIC/ (WxDxH) :25x25x50 LBS: 12 SPACE: FLOOR CM:63.5x63.5x127.0 KG: 5.40 <b>ELECT:</b> 120 VOLT/60 HZ BATTERY CHARGER 10.8-11.1V LITHIUM ION BATTERY - FOR AMBULATION	PHILIPS MEDICAL SYSTEMS 863283/989803176601 SURE SIGNS VS4	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: DICTATION  
ROOM NUMBER: 7027

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	096814 CRT03	1	<b>CART, LINEN</b> SS;3 FIXED SHELF;6" CASTERS (WxDxH) :62x24x70 LBS: 125 SPACE: FLOOR CM:157.5x61.0x177.8 KG: 56.70	G.S. MANUFACTURING GSM-G160 -	N	OF/OI
2	078147 CAB01	1	<b>CABINET, WARMING</b> MOBILE;2 COMPARTMENTS;20.6CF;GLASS DOORS;SEPARATE (WxDxH) :32x25x74 LBS: 358 SPACE: FLOOR CM:81.3x63.5x188.0 KG: 162.40 <b>ELECT:</b> 120V;50/60HZ;1PH;15A;1.8KW NEMA 5-20P;20A-125V PLUG HOSPITAL GRADE SPECIFY DOOR HINGE;RIGHT OR LEFT - PLACEHOLDER FOR FULL SIZE, SINGLE COMPARTMENT	PEDIGO PRODUCTS, INC. P-2055 P-2055	N	OF/OI
3	094909 CRT07	2	<b>CART, UTILITY</b> MULTIPURPOSE CART;800LB CAPACITY;5"WHEELS (WxDxH) :21x31x42 LBS: 44 SPACE: FLOOR CM:53.3x78.7x106.7 KG: 20.00	VERSACART EZTOTE 390 101-390	E	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: CLEAN SUPPLY  
ROOM NUMBER: 7028

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	506535 OSS01	1	<b>OPEN SUPPLY SYSTEM</b>	ZZZ - GENERIC DESCRIPTIONS	N	OF/VI
			- PAREX SUPPLY SYSTEM, WEIGHT BASED -PROVIDE BACKING ON ALL WALLS WITH NO OUTLETS EXTENDING BEYOND 12" AFF. REQUIRES POWER AND DATA AT CONTROL LOCATION.			
2	077862 BIN01	30	<b>BIN, ACCESSORY, PANEL, LOUVERED</b> LOUVERED PANEL;36";STONE (WxDxH) :36x1x19 CM:91.4x2.5x48.3	PAR EXCELLENCE SYSTEMS, INC. 30-636	N	OF/CI
			SPACE: WALL			
3	084658 CRT01	3	<b>CART, CYLINDER, MEDICAL GAS</b> 6 CAPACITY; E/D SIZE;W/HEAVY DUTY 5" CASTERS (WxDxH) :11x16x39 CM:27.9x40.6x99.1	W.T. FARLEY, INC. CR-DC06E-G -	I	OF/OI
			LBS: 22 KG: 10.00			
			SPACE: FLOOR			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, CORRIDOR  
ROOM NUMBER: 7033

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	004335 CRT28	1	<b>CART, UTILITY</b> 2 SHELF;400 LB CAP;MOLDED PLASTIC;BEIGE (WxDxH) :18x39x34      LBS: 31      SPACE: FLOOR CM:45.7x99.1x86.4      KG: 14.10	RUBBERMAID COMM. PRODUCTS 4500-88BEIG 4500-88	E	OF/OI
2	096814 CRT03	1	<b>CART, LINEN</b> SS;3 FIXED SHELF;6" CASTERS (WxDxH) :62x24x70      LBS: 125      SPACE: FLOOR CM:157.5x61.0x177.8      KG: 56.70 - GOWN	G.S. MANUFACTURING GSM-G160 -	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: ALCOVE, EQUIPMENT  
ROOM NUMBER: 7035

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	089251 CRT05	1	<b>CART, RESUSCITATION, CARDIAC</b> 4 DRAWER;2 3",1 6",1 9";ADJ DEFIB TRAY;IV POLE (WxDxH) :44x24x52 CM:111.8x61.0x132.1 <b>ELECT:</b> 120V FOR POWER STRIP	INTERMETRO INDUSTRIES CORP. LECCRP3 LIFELINE	E	OF/OI
-DIMENSIONS LISTED DO NOT INCLUDE SIDE ACCESSORIES, IV POLE HEIGHT, OR DEFIBRILLATOR SHELF HEIGHT						
2	087848 DEF01	1	<b>DEFIBRILLATOR</b> GUIDELINES 2015 - COMPATIBLE, AED W/MANUAL OVERRIDE (WxDxH) :11x13x8 CM:27.9x33.0x20.3 <b>ELECT:</b> 120VAC;60HZ;2AMPS;SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER	ZOLL MEDICAL CORPORATION 30120000001110012 R SERIES ALS	E	OF/OI
3	092325	1	<b>ASPIRATOR</b> VACUUM;5 SETTINGS;30LPM;BUILT IN AC/DC POWER;LED (WxDxH) :13x7x13 CM:33.0x17.8x33.0 <b>ELECT:</b> 100-240VAC;50/60HZ;RECHARGEABLE 12VDC BATTERY	LAERDAL MEDICAL CORP. 78002001/ACC LSU W/CANISTER	E	OF/OI
4	079188 SHV01	1	<b>SHELVING, WIRE</b> 5 SHELVES;STATIONARY;CHROME;SHELF INLAY (WxDxH) :24x24x74 CM:61.0x61.0x188.0	INTERMETRO INDUSTRIES CORP. A2424NC/74P A2424NC	E	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: SERVICE ELEVATOR LOBBY  
ROOM NUMBER: 7084

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	063113	3	<b>STRETCHER</b>	STRYKER CORP/MEDICAL DIV	I	OF/OI
	STR01		700LB CAPACITY;3"MATTRESS;ZOOM DRIVE;30" WIDE; (WxDxH) :38x85x34 CM:96.5x215.9x86.4	PRIME SERIES 1125 ZOOM		
			SPACE: FLOOR			
			- FOR SPACE AND UTILITY PLANNING ONLY			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 7701

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	006264	1	<b>NOTE:</b>	ZZZ - NOTE: (U.S.)	E	OF/OI
			- 18 REGULAR ROOMS, 1 ISOLATION, 1 POSITIVE PRESSURE			
2	086193 BED01	20	<b>BED, MED-SURG</b> 500# CAP;IBED UPGRADEABLE;INBED SCALE;ISOFLEX (WxDxH) :43x93x30 SPACE: FLOOR CM:109.2x236.2x76.2	STRYKER CORP/MEDICAL DIV S3 PX2-3005 S3	E	OF/OI
			- CONFIRM O2 TANK HOLDER QUANTITIES			
3	053630	20	<b>BED ACCESSORIES</b> FULLY DYNAMIC;NON-PWR;PRESSURE MANAGEMENT SURFACE (WxDxH) :35x84x7 LBS: 50 SPACE: SPACE CM:88.9x213.4x17.8 KG: 22.70	B.G. INDUSTRIES ACCUMAX QUANTUM A3-117610-384	E	OF/OI
			<b>ELECT:</b> 110 VOLT/PHASE 1			
4	080568	20	<b>ALARM</b> ALERTS STAFF TO ATTEMPTED BED, CHAIR OR TOILET (WxDxH) :4x6x3 LBS: 1 SPACE: WALL CM:10.2x15.2x7.6 KG: 0.50	POSEY CO. 8345 SITTER ELITE	E	OF/OI
			<b>ELECT:</b> 4.6-6.3 VDC, 4 "AA" ALKALINE BATTERIES			
5	096840 LFT05	20	<b>LIFTER, PATIENT, ACCESSORY, RAIL</b> 10'STRAIGHT;FOR USE WITH C1000 MOTOR SPACE: CEILING	PRISM MEDICAL 10' STRAIGHT	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 7701

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
6	087162 MON02	10	<b>MONITOR, PHYSIOLOGICAL</b> 15" TOUCH SCREEN;4 WAVE;ICU S/W;MIB INTERFACE (WxDxH) :13x8x12 LBS: 13 SPACE: WALL CM:33.0x20.3x30.5 KG: 5.90 <b>ELECT:</b> 100-240V, 50/60HZ <b>PLUMB:</b> NO PLUMBING REQUIRED - *ARE IN THE PROCESS OF BEING PURCHASED AND WILL RELOCATE WITH ONCOLOGY.	PHILIPS MEDICAL SYSTEMS MX550 866066	E	OF/OI
7	077835 MNT01	20	<b>MONITOR ACCESSORY, MOUNTING ASSEMBLY</b> VARIABLE HEIGHT MOUNT;TILT/SWIVEL ADJUSTMENT; SPACE: WALL - *ARE IN THE PROCESS OF BEING PURCHASED AND WILL RELOCATE WITH ONCOLOGY.	GCX CORPORATION WA-0011-27 VHM-25	E	OF/OI
8	086984 MON06	10	<b>MONITOR, VITAL SIGNS</b> 8.4"LCD SCREEN;NBP;SPO2;TYMPANIC TEMP;WIRELESS (WxDxH) :11x6x9 LBS: 7 SPACE: WALL CM:27.9x15.2x22.9 KG: 3.10 <b>ELECT:</b> 120 VOLT/60 HZ BATTERY CHARGER LITHIUM ION BATTERY 10.8-11.1V DATA OUTPUT:HL7 FORMAT, VIA ETHERNET PORT SERIAL DATA - *WILL RELOCATE FROM ONCOLOGY	PHILIPS MEDICAL SYSTEMS 863283 SURESIGNS VS4	E	OF/OI
9	044072	20	<b>SPHYGMOMANOMETER, ANEROID, WALL MOUNT</b> ADULT;BUILT-IN BASKET;CHARCOAL;(1) ADULT BP CUFF (WxDxH) :7x6x8 LBS: 2 SPACE: WALL CM:17.8x15.2x20.3 KG: 0.90	WELCH ALLYN, INC. 7670-01 TYCOS 767	E	OF/OI





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 7701

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
10	093331 WCH01	20	<b>WALL CHANNEL</b> WALL CHANNEL;SEISMIC;19"  SPACE: WALL	PHILIPS MEDICAL SYSTEMS MXU175	N	OF/CI
11	028803	20	<b>STAND, MAYO</b> MOBILE;CHROME;HAND HEIGHT ADJUSTABLE 30-40 IN (WxDxH) :13x20x40      LBS: 22      SPACE: FLOOR CM:33.0x50.8x101.6      KG: 10.00 - PLACEHOLDER. SPECS TO BE DETERMINED. MAY NOT NEED 1 PER ROOM.	PEDIGO PRODUCTS, INC. P-65 MOBILE	E	OF/OI
12	067194	20	<b>I.V. POLE</b> STAINLESS STEEL;5 LEG;3 IN CASTERS;4 HOOK TOP (WxDxH) :23x6x35      LBS: 32      SPACE: FLOOR CM:58.4x15.2x88.9      KG: 14.50 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	PRYOR PRODUCTS 135	E	OF/OI
13	088643	20	<b>I.V. INFUSION PUMP</b> GENERAL INFUSION PUMP;SINGLE CHANNEL;POLE CLAMP; (WxDxH) :8x8x6      LBS: 10      SPACE: SPACE CM:20.3x20.3x15.2      KG: 4.50 <b>ELECT:</b> 120V;50/60HZ;35W;1A;1 RECHARGEABLE BATTERY LEAD-ACID 6V;POWER CORD:HOSPITAL-GRADE AC CORD 10 FT LONG W/TRANSPARENT PLUG AND RETAINER PLATE - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	HOSPIRA, INC. 30010 PLUM 360/MEDNET	E	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 7701

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
14	096809	20	<b>PUMP, COMPRESSION</b> DVT;LCD SCREEN;ONE BUTTON START;CARRY HANDLE (WxDxH) :9x8x10 LBS: 9 SPACE: SPACE CM:22.9x20.3x25.4 KG: 4.10 <b>ELECT:</b> 120V/60HZ - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	ARJOHUNTLEIGH ACS900 FLOWTRON	L	OF/OI
15	066897	20	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	MEDLINE INDUSTRIES, INC. MDS89668XW	E	OF/OI
16	087247	20	<b>WALKER, FOLDING, ADULT</b> SIDE WALKER;ALUMINUM;USER WT CAP 300LBS;HT ADJUST (WxDxH) :19x16x36 LBS: 4 SPACE: FLOOR CM:48.3x40.6x91.4 KG: 1.60 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	PERFORMANCE HEALTHCARE 081561752 HEMI DAYS INVACARE	E	OF/OI
17	063659	20	<b>TABLE, OVERBED</b> SINGLE TOP;W/VANITY;LOW PROFILE U-SHAPED BASE (WxDxH) :21x32x45 LBS: 53 SPACE: FLOOR CM:53.3x81.3x114.3 KG: 24.10 SPECIFY FINISH SPECIFY FINISH: BASE SPECIFY FINISH: EDGE	STRYKER CORP/MEDICAL DIV 3150-000-200 TRU-FIT	E	OF/OI





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-1 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 20)  
ROOM NUMBER: 7701

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
24	058768 HMP01	20	<b>HAMPER, LINEN</b> W/LID;FOOT PEDAL;STEEL FRAME;POWDER COAT FINISH (WxDxH) :21x22x33 SPACE: FLOOR CM:53.3x55.9x83.8 - INSTALLED IN CASEWORK	CENTURION MEDICAL PRODUCTS CX302 MEDIUM	E	OF/OI
25	078043 HMP02	20	<b>HAMPER, LINEN</b> W/LID;FOOT PEDAL OPERATED;TUBULAR CHROME FRAME; (WxDxH) :19x20x26 SPACE: FLOOR CM:48.3x50.8x66.0 - FOR YELLOW BAG CHEMO. INSTALLED IN CASEWORK	CENTURION MEDICAL PRODUCTS CENTURION CX502	E	OF/OI
26	047636	40	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60 - (1) FOR PATIENT ROOM, (1) FOR TOILET	RUBBERMAID COMM. PRODUCTS 2544 40 QT	E	OF/OI
27	085429 RLS01	40	<b>RAIL SYSTEM</b> 7"2" VERTICAL RAIL SYSTEM;SURFACE MOUNT;1 TIER SPACE: WALL  - WITH BASKET, SHELF, AND OTHER ACCESSORIES TO BE DETERMINED.	MODULAR SERVICES COMPANY 530-4522	N	CF/CI
28	504469	20	<b>CLOCK</b>  - SPECIFIED BY OTHERS	ZZZ - GENERIC DESCRIPTIONS	N	OF/OI

END OF ROOM  
END OF DEPARTMENT



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: STAFF LOUNGE  
ROOM NUMBER: 7049

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	058112	2	<b>REFRIGERATOR/FREEZER, UPRIGHT</b> REF02 115V;18.2 CF;TOP FREEZER;NON-ICE;LEFT HINGE;WHITE (WxDxH) :30x32x67 LBS: 205 SPACE: FLOOR CM:76.2x81.3x170.2 KG: 93.00 ELECT: 115V;4.5A;60HZ;407 WATTS	SEARS COMMERCIAL SALES KENMORE 46-68972	N	OF/OI
2	080483	2	<b>OVEN, MICROWAVE</b> OVE01 1000W;S.S. INTERIOR/EXTERIOR (WxDxH) :21x16x13 LBS: 37 SPACE: COUNTER CM:53.3x40.6x33.0 KG: 16.80 ELECT: 120V;14A/11;1.6KW;15AMP NEMA 15-R RECEPTACLE	SHARP ELECTRONICS CORP. R21LCF R21LCF	N	OF/OI
3	090442	1	<b>COFFEE BREWER</b> COF02 SINGLE-CUP BREWER FOR COFFEE, TEA, HOT CHOCOLATE (WxDxH) :11x14x14 LBS: 18 SPACE: COUNTER CM:27.9x35.6x35.6 KG: 8.20 ELECT: 120VAC, 1400 WATTS - TO INCLUDE HOT WATER TAP	KEURIG, INC. COMMERCIAL BREWING K150	L	OF/VI
4	051789	1	<b>TOASTER, COMMERCIAL</b> TST01 2-SLOT;CHROME-PLATED CONSTRUCTION;1 3/8"WIDE SLOTS (WxDxH) :14x9x9 LBS: 5 SPACE: COUNTER CM:35.6x22.9x22.9 KG: 2.30 ELECT: 120 VOLTS/8 AMPS/925 WATTS/PHASE 1	WARING PRODUCTS WCT702 -	N	OF/OI
5	006677	2	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 LBS: 43 SPACE: FLOOR CM:30.5x53.3x88.9 KG: 19.50 SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: STAFF LOUNGE  
ROOM NUMBER: 7049

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
6	079040	1	<b>CHAIR, MASSAGE</b> 6 MOTOR;2 AIR PUMPS;FULL BODY;RECLINER;LCD SCREEN (WxDxH) :31x53x49 LBS: 220 SPACE: FLOOR CM:78.7x134.6x124.5 KG: 99.80 <b>ELECT:</b> AC 120V;60HZ;20-230W	BEAUTYHEALTH BC-07DH	N	OF/OI

END OF ROOM



A division of **Ross&Baruzzini**

712A

**MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT**

-

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: CORRIDOR  
ROOM NUMBER: 7051

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	093412 MVD01	4	<b>MONITOR, VIDEO</b> MXU0299: DISPLAY FLAT PANEL NON-TOUCH PIIC SPACE: SPACE	PHILIPS MEDICAL SYSTEMS MXU0299 866390	N	OF/OI
2	093331 WCH01	4	<b>WALL CHANNEL</b> WALL CHANNEL;SEISMIC;19" SPACE: WALL	PHILIPS MEDICAL SYSTEMS MXU175	N	OF/CI

**END OF ROOM**



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: PATIENT CARE STATION (QTY 5)  
ROOM NUMBER: 7052

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	087769 MON01	2	<b>MONITOR, CENTRAL STATION</b> PIIC IX SURVEILLANCE CENTRAL MONITORING;19" TOUCH (WxDxH) :48x16x16 CM:121.9x40.6x40.6 SPACE: COUNTER <b>ELECT:</b> CPU: 115 VAC;60 HZ;5.2 AMPS LCD DISPLAY: 115 VAC;60 HZ;48W TYP. UPS: 115 VAC;60 HZ;5.6 AMPS PRINTER: 117 VAC;60 HZ;5.3 AMPS <b>HVAC:</b> MUST MAINTAIN TEMP AT 62-82 DEGREES F AND NON-CONDENSING RELATIVE HUMIDITY AT 30-60% CPU: 215 BTU/HR UPS: 437 BTU/HR (MAX) DATA CONNECTION REQUIRED, SEE TECH SHEET	PHILIPS MEDICAL SYSTEMS INTELLIVUE INFO CENTER IX	N	OF/OI
2	505366	1	<b>MONITORING SYSTEM</b>	ZZZ - MONITORING	N	OF/OI
			- ALLOWANCE FOR MONITORING INFRASTRUCTURE			
3	086723 ANA01	2	<b>LAB ANALYZER, GLUCOSE</b> BLOOD GLUCOSE & KETONE MONITORING;BARCODE READER (WxDxH) :3x2x8 CM:7.6x5.1x20.3 LBS: 1 KG: 0.50 SPACE: SPACE <b>ELECT:</b> 2 AA BATTERIES (ALKALINE, LITHIUM, OR NICAD)	ABBOTT DIAGNOSTICS 71411-70 PRECISION PRO	N	OF/OI

END OF ROOM





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: ALCOVE, CART  
ROOM NUMBER: 7058

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - IV CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	I	OF/OI
2	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - RT CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	I	OF/OI
3	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - FOR LINEN	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	I	OF/OI
4	092750 MON05	1	<b>MONITOR, VITAL SIGNS</b> NIBP;SPO2;TEMPERATURE;RECORDER;ADULT/PEDIATRIC/ (WxDxH) :25x25x50 LBS: 12 SPACE: FLOOR CM:63.5x63.5x127.0 KG: 5.40 <b>ELECT:</b> 120 VOLT/60 HZ BATTERY CHARGER 10.8-11.1V LITHIUM ION BATTERY - FOR AMBULATION	PHILIPS MEDICAL SYSTEMS 863283/989803176601 SURE SIGNS VS4	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: ALCOVE, CART  
ROOM NUMBER: 7058

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
5	089251 CRT05	1	<b>CART, RESUSCITATION, CARDIAC</b> 4 DRAWER;2 3",1 6",1 9";ADJ DEFIB TRAY;IV POLE (WxDxH) :44x24x52 SPACE: FLOOR CM:111.8x61.0x132.1 <b>ELECT:</b> 120V FOR POWER STRIP	INTERMETRO INDUSTRIES CORP. LECCRP3 LIFELINE	E	OF/OI
-DIMENSIONS LISTED DO NOT INCLUDE SIDE ACCESSORIES, IV POLE HEIGHT, OR DEFIBRILLATOR SHELF HEIGHT						
6	087848 DEF01	1	<b>DEFIBRILLATOR</b> GUIDELINES 2015 - COMPATIBLE, AED W/MANUAL OVERRIDE (WxDxH) :11x13x8 LBS: 14 SPACE: COUNTER CM:27.9x33.0x20.3 KG: 6.40 <b>ELECT:</b> 120VAC;60HZ;2AMPS;SEALED LEAD ACID BATTERY INTEGRAL BATTERY CHARGER	ZOLL MEDICAL CORPORATION 30120000001110012 R SERIES ALS	N	OF/OI
7	092325	1	<b>ASPIRATOR</b> VACUUM;5 SETTINGS;30LPM;BUILT IN AC/DC POWER;LED (WxDxH) :13x7x13 LBS: 9 SPACE: SPACE CM:33.0x17.8x33.0 KG: 4.10 <b>ELECT:</b> 100-240VAC;50/60HZ;RECHARGEABLE 12VDC BATTERY	LAERDAL MEDICAL CORP. 78002001/ACC LSU W/CANISTER	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: MEDICATION  
ROOM NUMBER: 7059

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	072239	1	<b>DISPENSER, MEDICATION, AUTOMATED</b> 14.75"LCD MONITOR;KEYBOARD;TOUCHPAD;BIO-ID SCANNER (WxDxH) :23x27x55 LBS: 166 SPACE: FLOOR CM:58.4x68.6x139.7 KG: 75.30 <b>ELECT:</b> 120V;60HZ;3 AMPS MAX EMERGENCY POWER REQUIRED <b>HVAC:</b> 409 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET - NEED TO UPDATE TO ES SERIES	BD CAREFUSION PYXIS MEDSTATION ES 6 DRAWER MAIN	N	OF/VI
2	072241	1	<b>DISPENSER, MEDICATION, AUTOMATED</b> SLIDE BRACKETS W/DRAWER RELEASE;PLASTIC TOP (WxDxH) :23x47x47 LBS: 133 SPACE: FLOOR CM:58.4x119.4x119.4 KG: 60.30 <b>ELECT:</b> 120V;60HZ;3 AMPS MAX;EMERGENCY POWER REQUIRED <b>HVAC:</b> 222 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET	BD CAREFUSION PYXIS 7 DRAWER AUXILIARY MEDSTATION ES	N	OF/OI
3	072243	1	<b>DISPENSER, MEDICATION, AUTOMATED</b> SINGLE COLUMN;1 STORAGE SHELF PER DOOR;INTERIOR (WxDxH) :31x28x80 LBS: 314 SPACE: FLOOR CM:78.7x71.1x203.2 KG: 142.40 <b>ELECT:</b> 120V;60HZ;3 AMPS;EMERGENCY POWER REQUIRED <b>HVAC:</b> 222 BTU/HR DATA CONNECTION REQUIRED, SEE TECH SHEET	BD CAREFUSION PYXIS FOUR DOOR AUXILIARY MEDSTATION ES	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: MEDICATION  
ROOM NUMBER: 7059

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
4	079417 REF01	1	<b>REFRIGERATOR, UPRIGHT, MEDICAL GRADE</b> 115V;SINGLE DOOR;19.7 CF;SS;LED DIG. TEMP;LOCKING (WxDxH) :30x31x80 LBS: 395 SPACE: FLOOR CM:76.2x78.7x203.2 KG: 179.20 <b>ELECT:</b> 115V,60HZ,8.7A <b>HVAC:</b> 950 BTU/HR MAX HEAT REJECTION	FOLLETT CORPORATION REF20-LB-R000 REF20-LB-R000	N	OF/OI
5	053907 MED01	1	<b>DISPENSER, MEDICATION, ACCESSORY</b> MED DISPENSING UNIT REFRIGERATOR CONTROL (WxDxH) :3x6x7 LBS: 4 SPACE: SPACE CM:7.6x15.2x17.8 KG: 1.80 <b>ELECT:</b> POWERED FROM MAIN DISPENSER	BD CAREFUSION PYXIS REMOTE MANAGER MSRM	N	OF/VI
6	506535 OSS01	1	<b>OPEN SUPPLY SYSTEM</b>  - PAREX SUPPLY SYSTEM, WEIGHT BASED -PROVIDE BACKING ON ALL WALLS WITH NO OUTLETS EXTENDING BEYOND 12" AFF. REQUIRES POWER AND DATA AT CONTROL LOCATION.	ZZZ - GENERIC DESCRIPTIONS	N	OF/VI
7	077862 BIN01	2	<b>BIN, ACCESSORY, PANEL, LOUVERED</b> LOUVERED PANEL;36";STONE (WxDxH) :36x1x19 SPACE: WALL CM:91.4x2.5x48.3	PAR EXCELLENCE SYSTEMS, INC. 30-636	N	OF/CI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: MEDICATION  
ROOM NUMBER: 7059

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
8	093412 MVD01	1	<b>MONITOR, VIDEO</b> MXU0299: DISPLAY FLAT PANEL NON-TOUCH PIIC SPACE: SPACE	PHILIPS MEDICAL SYSTEMS MXU0299 866390	N	OF/OI
			- TO INCLUDE SPEAKERS			
9	055704 EWS01	1	<b>ENCLOSURE, WALL, SHARPS</b> 2 OR 5 QT WALL ENCLOSURE (WxDxH) :13x6x12 CM:33.0x15.2x30.5 SPACE: WALL	KENDALL HEALTHCARE, COVIDIEN 85161H IN-ROOM	L	OF/CI
10	093535 DSL02	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> 2GAL;BLACK;WHITE HINGED LID W/PORT;LEAK RESISTANT (WxDxH) :12x10x12 CM:30.5x25.4x30.5 SPACE: WALL	KENDALL HEALTHCARE, COVIDIEN 8602RC/8963 8602RC/8963	N	OF/OI
11	061957 GLV02	1	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 CM:45.7x10.2x22.9 SPACE: WALL	CLINTON INDUSTRIES G-1030 -	N	OF/CI
12	006677	1	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 CM:30.5x53.3x88.9 LBS: 43 KG: 19.50 SPACE: FLOOR SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: MEDICATION  
ROOM NUMBER: 7059

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
13	047636	2	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60 - PLACEHOLDER FOR RED AND BLACK BINS. NEED VENDOR AND MODEL.	RUBBERMAID COMM. PRODUCTS 2544 40 QT	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: CLEAN SUPPLY  
ROOM NUMBER: 7063

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	506535 OSS01	1	<b>OPEN SUPPLY SYSTEM</b>	ZZZ - GENERIC DESCRIPTIONS	N	OF/VI
			- PAREX SUPPLY SYSTEM, WEIGHT BASED -PROVIDE BACKING ON ALL WALLS WITH NO OUTLETS EXTENDING BEYOND 12" AFF. REQUIRES POWER AND DATA AT CONTROL LOCATION.			
2	077862 BIN01	30	<b>BIN, ACCESSORY, PANEL, LOUVERED</b> LOUVERED PANEL;36";STONE (WxDxH) :36x1x19 CM:91.4x2.5x48.3	PAR EXCELLENCE SYSTEMS, INC. 30-636	N	OF/CI
			SPACE: WALL			
3	084658 CRT01	3	<b>CART, CYLINDER, MEDICAL GAS</b> 6 CAPACITY; E/D SIZE;W/HEAVY DUTY 5" CASTERS (WxDxH) :11x16x39 CM:27.9x40.6x99.1	W.T. FARLEY, INC. CR-DC06E-G -	N	OF/OI
			LBS: 22 KG: 10.00			
			SPACE: FLOOR			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: SOILED HOLDING  
ROOM NUMBER: 7064

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	094910 COM01	1	<b>COMPACTOR, WASTE</b> 20:1 COMPACT RATION;COMMERCIAL;ELECTROHYDRAULIC (WxDxH) :43x31x90 LBS: 1210 SPACE: FLOOR CM:109.2x78.7x228.6 KG: 548.80 <b>ELECT:</b> 115V;60HZ;15 AMP;SINGLE PHASE;DEDICATED CIRCUIT - ECO-PACK	FOUNTAIN INDUSTRIES ECO-PACK 3600	N	OF/OI
2	079188 SHV01	1	<b>SHELVING, WIRE</b> 5 SHELVES;STATIONARY;CHROME;SHELF INLAY (WxDxH) :24x24x74 SPACE: FLOOR CM:61.0x61.0x188.0	INTERMETRO INDUSTRIES CORP. A2424NC/74P A2424NC	N	OF/OI
3	073358 TRK01	1	<b>TRUCK/BIN, LINEN</b> HORIZONTAL SHELVES;POLYETHYLENE;IMPACT RESISTANT (WxDxH) :48x29x64 SPACE: FLOOR CM:121.9x73.7x162.6	MEESE ORBITRON DUNNE CO. POLY TRUX 90P	N	OF/OI
4	066897	1	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0	MEDLINE INDUSTRIES, INC. MDS89668XW	N	OF/OI
5	096813	1	<b>WASTE RECEPTACLE, STEP-ON</b> 12 GAL;WHITE;FOOT PEDAL OPERATED;FIRE SAFE (WxDxH) :12x12x23 SPACE: FLOOR CM:30.5x30.5x58.4	RUBBERMAID COMM. PRODUCTS RCPST12ERBWHI DEFENDERS	N	OF/OI





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: SOILED HOLDING  
ROOM NUMBER: 7064

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
6	071200	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> 9 GALLON;CLEAR SLIDE TOP;RECYCLED PLASTIC-BLACK (WxDxH) :18x12x19 SPACE: SPACE CM:45.7x30.5x48.3 SPECIFY ACCESSORIES AS SEPARATE LINE ITEMS	BECTON DICKINSON RCRA HAZARDOUS WASTE 305069	N	OF/OI
7	070664	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> CHEMOTHERAPY;PLASTIC;18 GAL;YELLOW;SLIDING LID (WxDxH) :18x13x26 LBS: 8 SPACE: FLOOR CM:45.7x33.0x66.0 KG: 3.60	KENDALL HEALTHCARE, COVIDIEN 8939 CHEMO DISPOSAL	N	OF/OI
8	089948	1	<b>WASTE RECEPTACLE, RECYCLING</b> 23 GALLON;DARK BLUE;RECYCLE SYMBOL;W/ LID (WxDxH) :20x11x35 SPACE: FLOOR CM:50.8x27.9x88.9	RUBBERMAID COMM. PRODUCTS H-1385BLU SLIMJIM 3540-07	N	OF/OI
9	079178	1	<b>WASTE RECEPTACLE</b> CONFIDENTIAL;SIDE PAPER SLOT;KEYPAD LOCK;SHIPS (WxDxH) :11x24x25 LBS: 30 SPACE: FLOOR CM:27.9x61.0x63.5 KG: 13.60	RUBBERMAID COMM. PRODUCTS 9W25 SLIM JIM	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: ALCOVE, CART  
ROOM NUMBER: 7065

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - IV CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	I	OF/OI
2	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - RT CART, SPECS TBD.	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	I	OF/OI
3	092815 CRT04	1	<b>CART, PROCEDURE</b> 24" CART;5-DRWR: 3-3", 1-6", 1-9";BEIGE SHELL/ (WxDxH) :35x26x40 CM:88.9x66.0x101.6 SPACE: FLOOR - FOR LINEN	ARMSTRONG MEDICAL INDUSTRIES PBL-B-24 PREMIER	I	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: EQUIPMENT STORAGE  
ROOM NUMBER: 7066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	006264	1	<b>NOTE:</b>	ZZZ - NOTE: (U.S.)	I	OF/OI
			- PLACE HOLDER FOR EQUIPMENT			
2	004335 CRT28	2	<b>CART, UTILITY</b> 2 SHELF;400 LB CAP;MOLDED PLASTIC;BEIGE (WxDxH) :18x39x34 LBS: 31 SPACE: FLOOR CM:45.7x99.1x86.4 KG: 14.10	RUBBERMAID COMM. PRODUCTS 4500-88BEIG 4500-88	E	OF/OI
3	087247	6	<b>WALKER, FOLDING, ADULT</b> SIDE WALKER;ALUMINUM;USER WT CAP 300LBS;HT ADJUST (WxDxH) :19x16x36 LBS: 4 SPACE: FLOOR CM:48.3x40.6x91.4 KG: 1.60	PERFORMANCE HEALTHCARE 081561752 HEMI DAYS INVACARE	E	OF/OI
4	096810 BRD01	2	<b>BOARD, PATIENT TRANSFER</b> 529 LB CAPACITY;FOR USE IN RADIOLOGY,SURGERY,ICU (WxDxH) :20x70x2 LBS: 8 SPACE: WALL CM:50.8x177.8x5.1 KG: 3.80	SAMARIT 440.0100 HIGHTEC	E	OF/OI
			- PLACE HOLDER			
5	078455 BRD03	1	<b>BOARD, PATIENT TRANSFER</b> SURGIBOARD 47X16";TO AND FROM OR TABLE (WxDxH) :16x47x2 SPACE: WALL CM:40.6x119.4x5.1	MCAULEY MEDICAL INC 440.0400 SAMARIT	E	OF/OI
6	094909 CRT07	2	<b>CART, UTILITY</b> MULTIPURPOSE CART;800LB CAPACITY;5"WHEELS (WxDxH) :21x31x42 LBS: 44 SPACE: FLOOR CM:53.3x78.7x106.7 KG: 20.00	VERSACART EZTOTE 390 101-390	E	OF/OI
			- NEED TO CONFIRM EXISTING			



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: EQUIPMENT STORAGE  
ROOM NUMBER: 7066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
7	074251	3	<b>CRUTCHES</b> ALUMINUM;HAND GRIP;I-BEAM ADJUSTMENT;300LB CAP SPACE: SPACE	MEDLINE INDUSTRIES, INC. QUICK FIT MDS80540	E	OF/OI
8	091438	1	<b>CHAIR, SHOWER</b> 500LB USER WT CAP;14-19" HT ADJ SEAT;W/KNOCK DOWN (WxDxH) :18x20x30 LBS: 8 SPACE: FLOOR CM:45.7x50.8x76.2 KG: 3.40	DRIVE MEDICAL 12021KD-1 DELUXE BARI	E	OF/OI
9	500146 BED04	1	<b>BED, MED-SURG</b>  - CAMTEC-44800	ZZZ - GENERIC DESCRIPTIONS	E	OF/OI
10	086346 ULT01	1	<b>ULTRASOUND SYSTEM</b> BLADDER 3D ULTRASOUND SCANNER;0-999ML;MOBILE (WxDxH) :24x24x53 SPACE: FLOOR CM:61.0x61.0x134.6 <b>ELECT:</b> 100-240VAC;SINGLE PHASE;50-60HZ;1.4A ADAPTER RECHARGABLE LITHIUM ION BATTERY 24V DC BATTERY CHARGER	VERATHON MEDICAL BLADDERSCAN PRIME 0270-0870	E	OF/OI
11	046403 IMU01	2	<b>IMAGING, ULTRASOUND SCANNER</b> BLADDER ULTRASOUND SCANNER; W/ROLLING CART (WxDxH) :15x15x30 LBS: 20 SPACE: FLOOR CM:38.1x38.1x76.2 KG: 9.10 <b>ELECT:</b> BATTERY CHARGER:100-120V;60HZ;INPUT CURRENT 0.39A 7.2V NIMH BATTERY PACK;STANDARD ELECTRIC OUTLET	VERATHON MEDICAL BVI3000 BLADDERSCAN	E	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: EQUIPMENT STORAGE  
ROOM NUMBER: 7066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
12	070664	1	<b>DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL</b> CHEMOTHERAPY;PLASTIC;18 GAL;YELLOW;SLIDING LID (WxDxH) :18x13x26 LBS: 8 SPACE: FLOOR CM:45.7x33.0x66.0 KG: 3.60	KENDALL HEALTHCARE, COVIDIEN 8939 CHEMO DISPOSAL	E	OF/OI
13	097533 CRT31	10	<b>CART, SUPPLY</b> YELLOW;STORAGE/TRANSPORT;ROTOMOLDED-HDPE (WxDxH) :20x20x40 LBS: 38 SPACE: FLOOR CM:50.8x50.8x101.6 KG: 17.20 - CHEMO CART	KADDIES INC K-1 PORTA-NIZER	E	OF/OI
14	075890	2	<b>BALL, THERAPY</b> BALL;25.5" DIAMETER;INFLATION REQUIRED;LATEX FREE (WxDxH) :26x26x26 SPACE: SPACE CM:66.0x66.0x66.0	PERFORMANCE HEALTHCARE 812803 GYMNIC	E	OF/OI
15	016423	1	<b>STOOL, FOOT, WITH HANDRAIL</b> CHROME PLATED STEEL;SKID RESISTANT RUBBER MAT (WxDxH) :12x15x36 LBS: 12 SPACE: FLOOR CM:30.5x38.1x91.4 KG: 5.50	BREWER COMPANY 11220 -	E	OF/OI
16	003508 MIR01	1	<b>MIRROR, POSTURE</b> SINGLE;MOBILE;MOUNTED ON 2 IN. SWIVEL CASTERS (WxDxH) :24x18x72 LBS: 79 SPACE: FLOOR CM:61.0x45.7x182.9 KG: 35.90	HAUSMANN INDUSTRIES, INC. 1671 -	E	OF/OI
17	002702	1	<b>STOOL, REVOLVING</b> PNEUMATIC; WITHOUT BACK; WITH CASTERS (WxDxH) :15x15x18 LBS: 23 SPACE: FLOOR CM:38.1x38.1x45.7 KG: 10.50 SPECIFY COLOR: IF OTHER THAN STANDARD	RELIANCE MEDICAL,A HAAG STREIT CO. 5340 5300 SERIES	E	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: EQUIPMENT STORAGE  
ROOM NUMBER: 7066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
18	081983 MON26	4	<b>MONITOR, VITAL SIGNS</b> MOBILE;BLOOD PRESSURE, HEART RATE, AND MAP (WxDxH) :21x21x51 LBS: 6 SPACE: FLOOR CM:53.3x53.3x129.6 KG: 2.70 <b>ELECT:</b> 120V, 60HZ	WELCH ALLYN, INC. 53STP-E1/4700-60 300 SERIES	E	OF/OI
19	097005 WLK01	1	<b>WALKER</b> PVC;18"BETWEEN ARMS;ANTI-TIP OUTRIGGER;3"CASTERS (WxDxH) :31x36x46 SPACE: FLOOR CM:78.7x91.4x116.9	OCELCO 242-469 AMBULATOR	E	OF/OI
20	008102	1	<b>WASTE RECEPTACLE</b> PLASTIC;23 GAL;RECTANGULAR;VENTING CHANNELS;4/PACK (WxDxH) :20x11x30 LBS: 8 SPACE: FLOOR CM:50.8x27.9x76.2 KG: 3.60 SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540 SLIM JIM	E	OF/OI
21	014753 SHV02	1	<b>SHELVING, WIRE</b> MOBILE;CHROME;5 SHELF;4 CASTER STEM BRAKES (WxDxH) :51x25x69 LBS: 60 SPACE: FLOOR CM:129.6x63.5x175.3 KG: 27.30	INTERMETRO INDUSTRIES CORP. 2448NC/63UP/ACC. SUPER ERECTA	I	OF/VI
22	058498	1	<b>SPHYGMOMANOMETER, ANEROID, MOBILE</b> 5 LEG STAND;CERTIFIED ACCURATE TO 3MMHG;LATEX (WxDxH) :18x18x36 LBS: 12 SPACE: FLOOR CM:45.7x45.7x91.4 KG: 5.50	WELCH ALLYN, INC. 7670-03	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: EQUIPMENT STORAGE  
ROOM NUMBER: 7066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
23	072678	1	<b>DOPPLER, FLOW DETECTOR</b> NON-DIRECTIONAL POCKET DOPPLER W/BUILT-IN (WxDxH) :18x18x36 LBS: 2 SPACE: FLOOR CM:45.7x45.7x91.4 KG: 0.90 <b>ELECT:</b> 9V ALKALINE BATTERY - NEED TO CONFIRM EXISTING	HUNTLEIGH HEALTHCARE D900-P-USA/EZ8 D900 DOPPLEX	I	OF/OI
24	086490	1	<b>OPHTHALMOSCOPE/OTOSCOPE</b> MOBILE STAND; PANOPTIC OPHTHALMOSCOPE;MACROVIEW (WxDxH) :22x22x54 SPACE: FLOOR CM:55.9x55.9x137.2 <b>ELECT:</b> 100-240V;50/60HZ;.18 AMPS MAX - NEED TO CONFIRM EXISTING	WELCH ALLYN, INC. 77710-82M/ACC GS 777 SERIES	I	OF/OI
25	067194	2	<b>I.V. POLE</b> STAINLESS STEEL;5 LEG;3 IN CASTERS;4 HOOK TOP (WxDxH) :23x6x35 LBS: 32 SPACE: FLOOR CM:58.4x15.2x88.9 KG: 14.50 - NEED TO CONFIRM EXISTING	PRYOR PRODUCTS 135	I	OF/OI
26	088643	2	<b>I.V. INFUSION PUMP</b> GENERAL INFUSION PUMP;SINGLE CHANNEL;POLE CLAMP; (WxDxH) :8x8x6 LBS: 10 SPACE: SPACE CM:20.3x20.3x15.2 KG: 4.50 <b>ELECT:</b> 120V;50/60HZ;35W;1A;1 RECHARGEABLE BATTERY LEAD-ACID 6V;POWER CORD:HOSPITAL-GRADE AC CORD 10 FT LONG W/TRANSPARENT PLUG AND RETAINER PLATE - NEED TO CONFIRM EXISTING	HOSPIRA, INC. 30010 PLUM 360/MEDNET	I	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: EQUIPMENT STORAGE  
ROOM NUMBER: 7066

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
27	066897	2	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0	MEDLINE INDUSTRIES, INC. MDS89668XW	I	OF/OI
28	057580 LFT02	1	<b>CART</b> HOLDS HOVERJACK AIR SUPPLY AND LIFT (WxDxH) :25x18x38 LBS: 26 SPACE: FLOOR CM:63.5x45.7x96.5 KG: 11.80 - NEED TO CONFIRM EXISTING	HOVERTECH INTERNATIONAL HJC-100 HOVER JACK	I	OF/OI
29	055330	1	<b>LIFTER, PATIENT</b> INFLATABLE;NYLON OXFORD;TESTED UPT 1100LBS (WxDxH) :39x70x30 SPACE: FLOOR CM:99.1x177.8x76.2 <b>ELECT:</b> 120V/60HZ/1100 WATTS;GROUNDED;14' POWER CORD - NEED TO CONFIRM EXISTING	HOVERTECH INTERNATIONAL HJ3902/AIR400G/ACC HOVERJACK	I	OF/OI
30	073144	1	<b>TRANSILLUMINATOR</b> VEIN LOCATOR;INFRARED LIGHT;HAND HELD;BATTERY (WxDxH) :24x24x36 SPACE: FLOOR CM:61.0x61.0x91.4 - NEED TO CONFIRM EXISTING	BODIFLOW, LLC AV300 ACCUVEIN	I	OF/OI
31	082590	1	<b>WALKER</b> 4-WHEEL WALKER;8" NON-MARRING CASTERS;300LB CAP (WxDxH) :24x26x37 LBS: 20 SPACE: FLOOR CM:61.0x66.0x94.0 KG: 9.10 - NEED TO CONFIRM EXISTING	PERFORMANCE HEALTHCARE 081576974 ROLLATOR WALKER	I	OF/OI

END OF ROOM





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

-

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: ALCOVE, NUTRITION  
ROOM NUMBER: 7067

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	095453	1	<b>CART, FOOD</b>	LAKESIDE MFG. CO.	E	OF/OI
	CRT02		SS;6 TRAY CAPACITY;6"LEDGE SPACING;REMOVABLE DOOR (WxDxH) :33x22x46 LBS: 110 SPACE: FLOOR CM:83.8x55.9x116.9 KG: 49.90	SP-6373 ELITE		
			- SOILED TRAY CART			

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: ALCOVE, LINEN  
ROOM NUMBER: 7068

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	096814 CRT03	1	<b>CART, LINEN</b> SS;3 FIXED SHELF;6" CASTERS (WxDxH) :62x24x70 LBS: 125 SPACE: FLOOR CM:157.5x61.0x177.8 KG: 56.70 - PATIENT LINENS	G.S. MANUFACTURING GSM-G160 -	N	OF/OI
2	006698 WST02	1	<b>WASTE RECEPTACLE</b> 44 GALLON;ROUND W/FLAT LID;GRAY CONTAINER;DOLLY (WxDxH) :24x24x37 LBS: 24 SPACE: FLOOR CM:61.0x61.0x94.0 KG: 10.90 SPECIFY COLOR: LID - FOR CLEAN GOWNS	RUBBERMAID COMM. PRODUCTS 2643-60/2645-60/2640 BRUTE	N	OF/OI
3	078147 CAB01	1	<b>CABINET, WARMING</b> MOBILE;2 COMPARTMENTS;20.6CF;GLASS DOORS;SEPARATE (WxDxH) :32x25x74 LBS: 358 SPACE: FLOOR CM:81.3x63.5x188.0 KG: 162.40 <b>ELECT:</b> 120V;50/60HZ;1PH;15A;1.8KW NEMA 5-20P;20A-125V PLUG HOSPITAL GRADE SPECIFY DOOR HINGE;RIGHT OR LEFT - PLACEHOLDER FOR FULL SIZE, SINGLE COMPARTMENT	PEDIGO PRODUCTS, INC. P-2055 P-2055	N	OF/OI
4	002532 CRT06	1	<b>CART, UTILITY</b> 3 CHROME WIRE SHELVES;5" SWIVEL CASTERS;(2) ONE (WxDxH) :36x24x38 LBS: 39 SPACE: FLOOR CM:91.4x61.0x96.5 KG: 17.70	INTERMETRO INDUSTRIES CORP. 3SPN53DC 3SP SERIES	N	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 7074

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	058112	1	<b>REFRIGERATOR/FREEZER, UPRIGHT</b> 115V;18.2 CF;TOP FREEZER;NON-ICE;LEFT HINGE;WHITE (WxDxH) :30x32x67 LBS: 205 SPACE: FLOOR CM:76.2x81.3x170.2 KG: 93.00 <b>ELECT:</b> 115V;4.5A;60HZ;407 WATTS	SEARS COMMERCIAL SALES KENMORE 46-68972	N	OF/OI
2	087445	1	<b>ICE MAKER W/WATER DISPENSER</b> 12# STORAGE CAP;425#/24HR;LEVER;W/O FILTER (WxDxH) :17x24x33 LBS: 199 SPACE: COUNTER CM:43.2x61.0x83.8 KG: 90.30 <b>ELECT:</b> 115VAC/60HZ/1PH/11A/.8KW 8.5' CORD W/NEMA 5-15 HOSPITAL GRADE PLUG <b>HVAC:</b> 5000 BTU/HR  CLEARANCES: VENTILATION 6" TOP & RIGHT SIDES SERVICE 12" TOP <b>PLUMB:</b> REFER TO VENDOR TECHNICAL INFORMATION	FOLLETT CORPORATION 12CI425A-L AIR COOLED	N	OF/CI
3	080483	1	<b>OVEN, MICROWAVE</b> 1000W;S.S. INTERIOR/EXTERIOR (WxDxH) :21x16x13 LBS: 37 SPACE: COUNTER CM:53.3x40.6x33.0 KG: 16.80 <b>ELECT:</b> 120V;14A/11;1.6KW;15AMP NEMA 15-R RECEPTACLE	SHARP ELECTRONICS CORP. R21LCF R21LCF	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 7074

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
4	062450	1	<b>COFFEE BREWER</b> PLUMBED;1 LOWER BREWER/2 UPPER WARMERS;200 OZ CAP. (WxDxH) :9x22x19 LBS: 30 SPACE: COUNTER CM:22.9x55.9x48.3 KG: 13.60 <b>ELECT:</b> 120V/15AMPS;TANK HEATER 1425W;TOTAL WATTS 1800 CORD ATTACHED;REQ.2-WIRES PLUS GROUND SERVICE RATED 120V,SINGLE PHASE,60HZ. <b>PLUMB:</b> 20-90 PSI (138-621 KPA).MACHINES SUPPLIED W/ 1/4" MALE FLARE FITTING - TO INCLUDE HOT WATER TAP	BUNN-O-MATIC CORP. AXIOM-15-3-0000/ACC 38700.0000	L	OF/VI
5	051789	1	<b>TOASTER, COMMERCIAL</b> 2-SLOT;CHROME-PLATED CONSTRUCTION;1 3/8"WIDE SLOTS (WxDxH) :14x9x9 LBS: 5 SPACE: COUNTER CM:35.6x22.9x22.9 KG: 2.30 <b>ELECT:</b> 120 VOLTS/8 AMPS/925 WATTS/PHASE 1	WARING PRODUCTS WCT702 -	N	OF/OI
6	094915	1	<b>DISPENSER, FOOD SERVICE, BEVERAGE</b> SOFT DRINK;4 POST MIXING DISPENSING VALVES;LEG KIT (WxDxH) :11x26x28 LBS: 145 SPACE: COUNTER CM:27.9x66.0x71.1 KG: 65.80 <b>ELECT:</b> 120V;60HZ;7 AMPS <b>PLUMB:</b> SS LABELED INLET SYRUP/WATER;3/8"BARB - COORDINATION REQUIRED FOR SYRUP, CO2, ETC UNDER COUNTER	LANCER CORP. CED 500	L	OF/OI
7	061957	1	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 SPACE: WALL CM:45.7x10.2x22.9	CLINTON INDUSTRIES G-1030 -	N	OF/CI



A division of **Ross&Baruzzini**

712A

**MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT**

-

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: NOURISHMENT  
ROOM NUMBER: 7074

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
8	006677	1	<b>WASTE RECEPTACLE</b> 23 GAL;PLASTIC;RECTANGLE;VENTING CHANNELS (WxDxH) :12x21x35 LBS: 43 SPACE: FLOOR CM:30.5x53.3x88.9 KG: 19.50 SPECIFY COLOR	RUBBERMAID COMM. PRODUCTS 3540/2673-60 SLIM JIM	N	OF/OI

**END OF ROOM**



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: STORAGE, RT TANK  
ROOM NUMBER: 7080

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	084658 CRT01	3	<b>CART, CYLINDER, MEDICAL GAS</b> 6 CAPACITY; E/D SIZE; W/HEAVY DUTY 5" CASTERS (WxDxH) :11x16x39      LBS: 22      SPACE: FLOOR CM:27.9x40.6x99.1      KG: 10.00	W.T. FARLEY, INC. CR-DC06E-G -	N	OF/OI
2	070870	4	<b>TRUCK, CYLINDER</b> SINGLE; E-SIZE; HEIGHT ADJUSTABLE (WxDxH) :12x12x42      SPACE: FLOOR CM:30.5x30.5x106.7	MEDLINE INDUSTRIES, INC. HCS53006	E	OF/OI
3	058617	2	<b>RACK, CYLINDER</b> HOLDS 6: M7, M9, C, D, OR E SIZE CYLINDERS; GREEN (WxDxH) :15x12x20      SPACE: FLOOR CM:38.1x30.5x50.8	ANTHONY WELDED PRODUCTS, INC. 09-6060 6060	E	OF/OI

END OF ROOM



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 7723

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
1	006264	1	<b>NOTE:</b>	ZZZ - NOTE: (U.S.)	E	OF/OI
			- 10 REGULAR ROOMS, 1 ISOLATION, 1 POSITIVE PRESSURE			
2	086193 BED01	12	<b>BED, MED-SURG</b> 500# CAP;IBED UPGRADEABLE;INBED SCALE;ISOFLEX (WxDxH) :43x93x30 SPACE: FLOOR CM:109.2x236.2x76.2	STRYKER CORP/MEDICAL DIV S3 PX2-3005 S3	E	OF/OI
			- CONFIRM O2 TANK HOLDER QUANTITIES			
3	053630	12	<b>BED ACCESSORIES</b> FULLY DYNAMIC;NON-PWR;PRESSURE MANAGEMENT SURFACE (WxDxH) :35x84x7 LBS: 50 SPACE: SPACE CM:88.9x213.4x17.8 KG: 22.70	B.G. INDUSTRIES ACCUMAX QUANTUM A3-117610-384	E	OF/OI
			<b>ELECT:</b> 110 VOLT/PHASE 1			
4	080568	12	<b>ALARM</b> ALERTS STAFF TO ATTEMPTED BED, CHAIR OR TOILET (WxDxH) :4x6x3 LBS: 1 SPACE: WALL CM:10.2x15.2x7.6 KG: 0.50	POSEY CO. 8345 SITTER ELITE	E	OF/OI
			<b>ELECT:</b> 4.6-6.3 VDC, 4 "AA" ALKALINE BATTERIES			
5	096840 LFT05	12	<b>LIFTER, PATIENT, ACCESSORY, RAIL</b> 10'STRAIGHT;FOR USE WITH C1000 MOTOR SPACE: CEILING	PRISM MEDICAL 10' STRAIGHT	N	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 7723

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
6	086984 MON06	12	<b>MONITOR, VITAL SIGNS</b> 8.4"LCD SCREEN;NBP;SPO2;TYMPANIC TEMP;WIRELESS (WxDxH) :11x6x9 LBS: 7 SPACE: WALL CM:27.9x15.2x22.9 KG: 3.10 <b>ELECT:</b> 120 VOLT/60 HZ BATTERY CHARGER LITHIUM ION BATTERY 10.8-11.1V DATA OUTPUT:HL7 FORMAT, VIA ETHERNET PORT SERIAL DATA - *WILL RELOCATE FROM ONCOLOGY	PHILIPS MEDICAL SYSTEMS 863283 SURESIGNS VS4	E	OF/OI
7	093331 WCH01	12	<b>WALL CHANNEL</b> WALL CHANNEL;SEISMIC;19"  SPACE: WALL	PHILIPS MEDICAL SYSTEMS MXU175	N	OF/CI
8	077835 MNT01	12	<b>MONITOR ACCESSORY, MOUNTING ASSEMBLY</b> VARIABLE HEIGHT MOUNT;TILT/SWIVEL ADJUSTMENT;  SPACE: WALL	GCX CORPORATION WA-0011-27 VHM-25	E	OF/OI
9	044072	12	<b>SPHYGMOMANOMETER, ANEROID, WALL MOUNT</b> ADULT;BUILT-IN BASKET;CHARCOAL;(1) ADULT BP CUFF (WxDxH) :7x6x8 LBS: 2 SPACE: WALL CM:17.8x15.2x20.3 KG: 0.90	WELCH ALLYN, INC. 7670-01 TYCOS 767	E	OF/OI
10	028803	12	<b>STAND, MAYO</b> MOBILE;CHROME;HAND HEIGHT ADJUSTABLE 30-40 IN (WxDxH) :13x20x40 LBS: 22 SPACE: FLOOR CM:33.0x50.8x101.6 KG: 10.00 - PLACEHOLDER. SPECS TO BE DETERMINED. MAY NOT NEED 1 PER ROOM.	PEDIGO PRODUCTS, INC. P-65 MOBILE	E	OF/OI





A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 7723

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
11	067194	12	<b>I.V. POLE</b> STAINLESS STEEL;5 LEG;3 IN CASTERS;4 HOOK TOP (WxDxH) :23x6x35 LBS: 32 SPACE: FLOOR CM:58.4x15.2x88.9 KG: 14.50 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	PRYOR PRODUCTS 135	E	OF/OI
12	088643	12	<b>I.V. INFUSION PUMP</b> GENERAL INFUSION PUMP;SINGLE CHANNEL;POLE CLAMP; (WxDxH) :8x8x6 LBS: 10 SPACE: SPACE CM:20.3x20.3x15.2 KG: 4.50 <b>ELECT:</b> 120V;50/60HZ;35W;1A;1 RECHARGEABLE BATTERY LEAD-ACID 6V;POWER CORD:HOSPITAL-GRADE AC CORD 10 FT LONG W/TRANSPARENT PLUG AND RETAINER PLATE - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	HOSPIRA, INC. 30010 PLUM 360/MEDNET	E	OF/OI
13	096809	12	<b>PUMP, COMPRESSION</b> DVT;LCD SCREEN;ONE BUTTON START;CARRY HANDLE (WxDxH) :9x8x10 LBS: 9 SPACE: SPACE CM:22.9x20.3x25.4 KG: 4.10 <b>ELECT:</b> 120V/60HZ - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	ARJOHUNTLEIGH ACS900 FLOWTRON	L	OF/OI
14	066897	12	<b>CHAIR, COMMODE</b> BARIATRIC;850 LBS CAP.;DROP ARM;28"SEAT;STEAL (WxDxH) :29x21x26 SPACE: FLOOR CM:73.7x53.3x66.0 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	MEDLINE INDUSTRIES, INC. MDS89668XW	E	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 7723

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
15	087247	12	<b>WALKER, FOLDING, ADULT</b> SIDE WALKER;ALUMINUM;USER WT CAP 300LBS;HT ADJUST (WxDxH) :19x16x36 LBS: 4 SPACE: FLOOR CM:48.3x40.6x91.4 KG: 1.60 - CENTRALIZED. PROVIDED ON LIST FOR SPACE AND UTILITY PLANNING ONLY.	PERFORMANCE HEALTHCARE 081561752 HEMI DAYS INVACARE	E	OF/OI
16	063659	12	<b>TABLE, OVERBED</b> SINGLE TOP;W/VANITY;LOW PROFILE U-SHAPED BASE (WxDxH) :21x32x45 LBS: 53 SPACE: FLOOR CM:53.3x81.3x114.3 KG: 24.10 SPECIFY FINISH SPECIFY FINISH: BASE SPECIFY FINISH: EDGE	STRYKER CORP/MEDICAL DIV 3150-000-200 TRU-FIT	E	OF/OI
17	080275	12	<b>REGULATOR, SUCTION, CONTINUOUS &amp; INTERMITTENT</b> 0-200MMHG W/SELECTOR VALVE (OFF,INT,CONT);CANISTER (WxDxH) :4x6x4 LBS: 1 SPACE: WALL CM:10.2x15.2x10.2 KG: 0.50	BOEHRINGER LABORATORIES, INC. 3804/1496/2469 PLATINUM SERIES	E	OF/OI
18	055048	12	<b>FLOWMETER, OXYGEN</b> W/DISS ADAPTOR  SPACE: SPACE  CONFIRM MEDICAL GAS CONNECTION	AMVEX, DIVISION OF OHIO MEDICAL FM15ODH -	E	OF/OI
19	062114 DSP01	12	<b>DISPENSER</b> EMEBAG DISPENSER (WxDxH) :7x7x7 SPACE: WALL CM:17.8x17.8x17.8	CENTURION MEDICAL PRODUCTS EMED200 EMEBAG	N	OF/CI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 7723

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
20	091196 DSP03	12	<b>DISPENSER, MASK</b> PLEXIGLAS DISPENSER FOR SURGICAL FACE MASK BOXES (WxDxH) :9x6x9 LBS: 2 SPACE: WALL CM:22.9x15.2x22.9 KG: 0.90	MEDLINE INDUSTRIES, INC. NONFM122 NONFM122	N	OF/OI
21	087029 DSL01	12	<b>DISPOSAL CONTAINER, SHARPS W/BRACKET</b> LOCKING WALL CABINET FOR SHARPS CONTAINERS;2 GAL (WxDxH) :15x6x14 SPACE: WALL CM:38.1x15.2x35.6	STERICYCLE OC-02-2004 BIO SYSTEMS	L	OF/CI
22	061957 GLV02	12	<b>DISPENSER, GLOVE</b> VERTICAL;TRIPLE;WIRE;POWER COATED FINISH (WxDxH) :18x4x9 SPACE: WALL CM:45.7x10.2x22.9	CLINTON INDUSTRIES G-1030 -	N	OF/CI
23	058768 HMP01	12	<b>HAMPER, LINEN</b> W/LID;FOOT PEDAL;STEEL FRAME;POWDER COAT FINISH (WxDxH) :21x22x33 SPACE: FLOOR CM:53.3x55.9x83.8 - INSTALLED IN CASEWORK	CENTURION MEDICAL PRODUCTS CX302 MEDIUM	E	OF/OI
24	078043 HMP02	12	<b>HAMPER, LINEN</b> W/LID;FOOT PEDAL OPERATED;TUBULAR CHROME FRAME; (WxDxH) :19x20x26 SPACE: FLOOR CM:48.3x50.8x66.0 - FOR YELLOW BAG CHEMO. INSTALLED IN CASEWORK	CENTURION MEDICAL PRODUCTS CENTURION CX502	E	OF/OI
25	047636	24	<b>WASTE RECEPTACLE</b> 40 QT;RECTANGULAR;FIRE RESISTANT;FIBERGLASS;BEIGE (WxDxH) :12x15x20 LBS: 8 SPACE: FLOOR CM:30.5x38.1x50.8 KG: 3.60 - (1) FOR PATIENT ROOM, (1) FOR TOILET	RUBBERMAID COMM. PRODUCTS 2544 40 QT	E	OF/OI



A division of Ross & Baruzzini

712A

MITCHELL PLANNING  
MAINE MEDICAL CENTER  
EQUIPMENT REPORT BY DEPARTMENT

DEPARTMENT: E7-2 ONCOLOGY  
ROOM NAME: PATIENT ROOM (QTY 12)  
ROOM NUMBER: 7723

LINE	ID #	QTY	GENERIC	MANUFACTURER	STATUS	RESP
26	085429 RLS01	24	<b>RAIL SYSTEM</b> 7"2" VERTICAL RAIL SYSTEM;SURFACE MOUNT;1 TIER SPACE: WALL	MODULAR SERVICES COMPANY 530-4522	N	CF/CI
			- WITH BASKET, SHELF, AND OTHER ACCESSORIES TO BE DETERMINED.			
27	504469	12	<b>CLOCK</b>	ZZZ - GENERIC DESCRIPTIONS	N	OF/OI
			- SPECIFIED BY OTHERS			

END OF ROOM  
END OF DEPARTMENT

END OF REPORT