MAGNET GANTRY MOUNTING PATTERN

B78-65

REV. DATE: 04/08/10

NOTE: DISTANCES TO WALL ARE TYPICAL ONLY FOR THOSE CONFIGURATIONS WHERE THE MAGNET IS ORIENTED SQUARE TO A FOUR WALL ROOM. IF THE MAGNET IS TO BE ROTATED OUT OF SQUARE, OR THE ROOM IS A COMPLEX SHAPE, CONTACT GE SITE PLANNING FOR ASSISTANCE.

DETAIL NOT TO SCALE

THIS SHEET IS PART OF THE DOCUMENT SET LISTED ON SHEET C1 AND SHOULD NOT BE SEPARATED

This drawing is based on Sketch No.: 13nef005

PIM R3

RQ – 133811

Reviewed for Code Compliance
Inspections Division
Approved with Conditions

Date: 05/16/13

Planner: Benjamin Wallace

W/S

GE Healthcare

Healthcare Project Implementation – Design

Milwaukee,

SHEET TITLE: STRUCTURAL DETAILS

MODALITY TYPE: OPTIMA MR430s

THIS PLAN IS SUBMITTED TO SUGGEST LOCATION OF GE HEALTHCARE EQUIPMENT AND ASSOCIATED APPARATUS, ELECTRICAL WIRING DETAILS AND ROOM ARRANGEMENTS. IN PREPARING THIS PLAN, EVERY EFFORT HAS BEEN MADE TO CONFORM DETAILS TO ACTUAL EQUIPMENT TO BE INSTALLED. IT IS NOT TO BE USED AS A SHOP OR ACTUAL CONSTRUCTION PURPOSES. HOWEVER, AND THE COMPANY CANNOT ACCEPT RESPONSIBILITY FOR ANY DAMAGES RESULTING THEREFROM.

PROJECT TITLE:

ORTHOPAEDIC ASSOCIATES OF PORTLAND

PORTLAND, MAINE

PROJECT	REVISION
130260	01

DATE:	22.Jan.13
DRAWN BY:	TMS
CHECKED BY:	PMM
GON NO:	4057872
GON DT:	16.JAN.13

REVISION HISTORY:

JGA – 07.Mar.13

CHECKED BY: PMM

SHEET

S2