ELECTRICAL DETAIL
HORIZONTAL WALL DUCT (TYPICAL)

ELEC–5
REV. DATE: 03/19/04

DUCT WIDTH		MINIMUM DIVIDERS REQUIRED
24"	610mm	2
18"	457mm	2
10"	254mm	2
6"	152mm	1
4"	102mm	1

DETAIL NOT TO SCALE

ELECTRICAL DETAIL
VERTICAL WALL DUCT (TYPICAL)

ELEC–6
REV. DATE: 03/19/04

DUCT WIDTH		MINIMUM DIVIDERS REQUIRED
24"	610mm	2
18"	457mm	2
10"	254mm	2
6"	152mm	1
4"	102mm	1

DETAIL NOT TO SCALE

ELECTRICAL DETAIL
SURFACE FLOOR DUCT (TYPICAL)

ELEC–14
REV. DATE: 4/02/04

DUCT WIDTH		MINIMUM DIVIDERS REQUIRED
24"	610mm	2
18"	457mm	2
10"	254mm	2
6"	152mm	1
4"	102mm	1

DETAIL NOT TO SCALE

ELECTRICAL DETAIL
BOX WITH COVERPLATE (TYPICAL)

ELEC–8
REV. DATE: 09/30/94

DETAIL NOT TO SCALE

This drawing is based on Sketch No.: 13nef005

RQ – 133811 PIM R3

THIS SHEET IS PART OF THE DOCUMENT SET LISTED ON SHEET C1 AND SHOULD NOT BE SEPARATED

Reviewed for Code Compliance
Inspections Division
Approved with Conditions
Date: 05/16/13
Planner: Benjamin Wallace

GE Healthcare
Healthcare Project Implementation – Design
Healthcare, Milwaukee, WIS

SHEET TITLE: ELECTRICAL DETAILS
MODALITY TYPE: OPTIMA MR430s

THIS PLAN IS SUBMITTED TO SUGGEST LOCATION OF GE HEALTHCARE EQUIPMENT AND ASSOCIATED APPURTENANCES ELECTRICAL WIRING, TRAILS, NOTE AND DIMENSIONS. IN PREPARING THIS PLAN, EVERY EFFORT HAS BEEN MADE TO CONFORM DETAILS TO ACTUAL EQUIPMENT EXPECTED TO BE INSTALLED. IT IS NOT TO BE USED FOR ACTUAL DISTRIBUTION PURPOSES. SUNDRY AND THE COMPANY CANNOT ACCEPT RESPONSIBILITY FOR ANY DAMAGES RESULTING THEREFROM.

PROJECT TITLE:
ORTHOPAEDIC ASSOCIATES OF PORTLAND
PORTLAND, MAINE

PROJECT	REVISION
130260	01

DATE: 22.Jan.13
DRAWN BY: TMS
CHECKED BY: PMM
GON NO: 4057872
GON DT: 16.JAN.13

REVISION HISTORY:
JGA – 07.Mar.13
CHECKED BY: PMM

SHEET
E3