EQUIPMENT DETAIL — PENETRATION PANEL

M30–15F
REV. DATE: 06.JUN.12

NOTE:
- ALL DIMENSIONS IN INCHES, BRACKETED [ ] DIMENSIONS IN MM
- NOT TO SCALE
- INDICATES CENTER OF GRAVITY
- 15" [381] INSTALLATION CLEARANCE IS ONLY REQUIRED WHEN USING INSTALLATION DOLLIES

EQUIPMENT DETAIL — SECONDARY PENETRATION WALL (SPW)

M30–15P
REV. DATE: 06.JUN.12

NOTE:
- ALL DIMENSIONS ARE IN INCHES. ALL [ ] BRACKETED DIMENSIONS ARE IN MILLIMETERS
- GRAPHICS NOT TO SCALE
- INDICATES CENTER OF GRAVITY

EQUIPMENT DETAIL — MAIN DISCONNECT

M17–15E
REV. DATE: 26.JUN.12

EQUIPMENT DETAIL — HEAT EXCHANGER CABINET

M30–15B
REV. DATE: 06.JUN.12

NOTE:
- ALL DIMENSIONS IN INCHES
- [BRACKETED] EQUALS MM
- DETAIL NOT TO SCALE

EQUIPMENT DETAIL — SUPPLY AND RETURN CONNECTIONS AT TOP OF HEC

M30–15D
REV. DATE: 11/20/10

SEE DETAIL M30–15B FOR HEAT EXCHANGER CABINET DETAIL

EQUIPMENT DETAIL — MAGNET MONITOR

M16–15C
REV. DATE: 04/24/07

NOTE:
- INDICATES CENTER OF GRAVITY
- INDICATES AIR FLOW

EQUIPMENT DETAIL — DC LIGHTING CONTROLLER

M20–15/
M20–15A
REV. DATE: 02/24/09

NOTE:
- ALL DIMENSIONS IN INCHES AND [MILLIMETERS]
- APPROXIMATE WEIGHTS
  CONTROL PANEL 155lbs (70kg)
  AUTOTRANSFORMER 80lbs (36.29kg)
- BTU/HR
  CONTROL PANEL 1024 (300W)
  AUTOTRANSFORMER 171 (50W)

EQUIPMENT DETAIL — MR MUSIC SYSTEM

M67–15
REV. DATE: 02/21/01

EQUIPMENT DETAIL — CONSOLE, HOST MONITOR & KEYBOARD

M30–150
REV. DATE: 05/22/08

NOTE:
- ALL DIMENSIONS ARE IN INCHES ALL BRACKETED [ ] DIMENSIONS ARE IN MILLIMETERS
- DETAIL NOT TO SCALE

EQUIPMENT DETAIL — OPERATOR WORKSPACE CABINET

M06–15D
REV. DATE: 02/24/09

NOTE:
- ALL DIMENSIONS IN INCHES
- [BRACKETED] EQUALS MM
- DETAIL NOT TO SCALE

EQUIPMENT DETAIL — PNEUMATIC PATIENT ALERT CONTROL BOX

M48–15
REV. DATE: 06/21/96

Reviewed for Code Compliance
Inspections Division
Approved with Conditions
Date: 05/16/13
Planner: Benjamin Wallace

GE Healthcare
Healthcare Project Implementation – Desi
Milwaukee,

SHEET TITLE: EQUIPMENT DETAILS
MODALITY TYPE: OPTIMA MR450w

THIS PLAN IS SUBMITTED TO SUGGEST LOCATION OF GE HEALTHCARE EQUIPMENT AND ASSOCIATED APPARATUS, ELECTRICAL WIRING DETAILS AND ROOM ARRANGEMENTS. IN PREPARING THIS PLAN, EVERY EFFORT HAS BEEN MADE TO CONFORM DETAILS TO ACTUAL EQUIPMENT EXPECTED TO BE INSTALLED. IT IS NOT TO BE USED FOR ACTUAL CONSTRUCTION PURPOSES. THE HEALTHCARE FACILITY OWNER CANNOT ASSUME RESPONSIBILITY FOR ANY DAMAGES RESULTING THEREFROM.

This drawing is based on Sketch No.: 13nef006

PIM R6

RQ – 132725

PROJECT TITLE:
ORTHOPAEDIC ASSOCIATES OF PORTLAND
PORTLAND, MAINE

PROJECT	REVISION
130261	00

DATE:	22.Jan.13
DRAWN BY:	TMS
CHECKED BY:	PMM
GON NO:	4057873
GON DT:	16.Jan.13

REVISION HISTORY:

SHEET
D2

THIS SHEET IS PART OF THE DOCUMENT SET LISTED ON SHEET C1 AND SHOULD NOT BE SEPARATED